

Connecticut Department of Public Safety Forensic Science Laboratory

278 Colony Street, Meriden, CT 06451

External DNA Audit Report on Compliance with the FBI Director's Quality Assurance Standards for Forensic DNA Testing Laboratories

Conducted on July 11-13, 2011

| | |
|-----------------|-------------------------|
| Lonnie Ginsberg | NFSTC Lead Auditor |
| Lisa Gefrides | NFSTC Technical Auditor |
| Jason Hodges | NFSTC Technical Auditor |
| Christie Smith | NFSTC Technical Auditor |
| Sylvia Thurmond | NFSTC Technical Auditor |



This audit was performed under Cooperative Agreement #2007-MU-BX-K008
with the
National Institute of Justice
and the
National Forensic Science Technology Center

**"This document is to be used for pre-decisional purposes only by the
laboratory audited and NDIS in determining compliance with these
standards".**

National Forensic Science Technology Center Inc
7881 114th Avenue N.
Largo FL 33773
Tel (727) 549-6067 Fax (727) 549-6070

THE FBI QUALITY ASSURANCE STANDARDS

AUDIT FOR

FORENSIC DNA TESTING LABORATORIES

IN ACCORDANCE WITH
THE QUALITY ASSURANCE STANDARDS
FOR
FORENSIC DNA TESTING LABORATORIES
EFFECTIVE JULY 1, 2009

An Audit of: Connecticut Department of Public Safety Forensic Science Laboratory

Dates of Audit: July 11-13, 2011

Auditor(s): Lonnie Ginsberg


(Name)



(Signature)

Lisa Gefrides

(Name)



(Signature)

Jason Hodges

(Name)



(Signature)

Christie Smith

(Name)



(Signature)

Sylvia Thurmond

(Name)



(Signature)

Last Updated: July, 21, 2010

Table of Contents

| | |
|---|----|
| CHECKLIST OF GENERAL LABORATORY INFORMATION | 04 |
| STANDARD 3 - QUALITY ASSURANCE PROGRAM | 05 |
| STANDARD 4 - ORGANIZATION AND MANAGEMENT | 08 |
| STANDARD 5 - PERSONNEL | 09 |
| STANDARD 6 - FACILITIES | 21 |
| STANDARD 7 - EVIDENCE CONTROL | 23 |
| STANDARD 8 - VALIDATION | 25 |
| STANDARD 9 - ANALYTICAL PROCEDURES | 29 |
| STANDARD 10 - EQUIPMENT CALIBRATION AND MAINTENANCE | 34 |
| STANDARD 11 - REPORTS | 36 |
| STANDARD 12 - REVIEW | 38 |
| STANDARD 13 - PROFICIENCY TESTING | 41 |
| STANDARD 14 - CORRECTIVE ACTION | 45 |
| STANDARD 15 - AUDITS | 46 |
| STANDARD 16 - SAFETY | 47 |
| STANDARD 17 -OUTSOURCING | 48 |
| Appendix A: Findings and Responses | 51 |
| Appendix C: Auditor Self-Certification | 84 |
| Appendix D: Personnel Qualifications | 93 |
| Appendix E: Approved Validations | 98 |

Checklist of General Laboratory Information

1. Name of Laboratory: Connecticut Department of Public Safety Forensic Science Laboratory
2. Federal / State / Regional / County / Local / Other: _____
Laboratory (Choose one)
3. Approximate Population Size Served: 3.5 million
4. Uses a Contract Laboratory: Yes No
Name of Contract Laboratory(ies): _____
5. NDIS Participant: Yes No
6. Applying for NDIS Participation: Yes No NA (Choose one)
7. Technologies Used: (Choose those that apply)
 STRs
 YSTRs
 MtDNA
 Other: _____
8. Number of staff:
DNA analysts: 18
DNA trainees: 0
DNA technicians: 2
Laboratory support personnel: 0
DNA technical leader: Carl Ladd* (took over as mtDNA TL 8/2009); Mary Beth Raffin (mtDNA TL Jan-Aug 2009)
On site: Yes No
Casework CODIS administrator: Michael Bourke
9. Last audit conducted on: December 2008 (Internal audits conducted in 2009 and 2010)
External Audit Internal Audit (Choose one)
10. Audit Document Discussion Used (Revision Date): July 2009

Standard 3. Quality Assurance Program

| | Yes | No | N/A |
|--|-------------------------------------|--------------------------|--------------------------|
| 3.1 For the DNA laboratory's quality assurance program: | <input checked="" type="checkbox"/> | <input type="checkbox"/> | <input type="checkbox"/> |
| a. Does the DNA laboratory have an established and maintained documented quality system that is appropriate to the testing activities? | <input checked="" type="checkbox"/> | <input type="checkbox"/> | <input type="checkbox"/> |
| b. Is the quality system equivalent to or more stringent than what is required by these Standards? | <input checked="" type="checkbox"/> | <input type="checkbox"/> | <input type="checkbox"/> |

[Click Here For Discussion](#)

Comment

| | Yes | No | N/A |
|---|-------------------------------------|-------------------------------------|--------------------------|
| 3.1.1 Is the quality system documented in a manual that includes or references the following elements: | <input type="checkbox"/> | <input checked="" type="checkbox"/> | <input type="checkbox"/> |
| 3.1.1.1 Goals and objectives? | <input checked="" type="checkbox"/> | <input type="checkbox"/> | <input type="checkbox"/> |
| 3.1.1.2 Organization and management? | <input checked="" type="checkbox"/> | <input type="checkbox"/> | <input type="checkbox"/> |
| 3.1.1.3 Personnel? | <input checked="" type="checkbox"/> | <input type="checkbox"/> | <input type="checkbox"/> |
| 3.1.1.4 Facilities? | <input type="checkbox"/> | <input checked="" type="checkbox"/> | <input type="checkbox"/> |
| 3.1.1.5 Evidence control? | <input checked="" type="checkbox"/> | <input type="checkbox"/> | <input type="checkbox"/> |
| 3.1.1.6 Validation? | <input checked="" type="checkbox"/> | <input type="checkbox"/> | <input type="checkbox"/> |
| 3.1.1.7 Analytical procedures? | <input checked="" type="checkbox"/> | <input type="checkbox"/> | <input type="checkbox"/> |
| 3.1.1.8 Equipment calibration and maintenance? | <input checked="" type="checkbox"/> | <input type="checkbox"/> | <input type="checkbox"/> |
| 3.1.1.9 Reports? | <input checked="" type="checkbox"/> | <input type="checkbox"/> | <input type="checkbox"/> |
| 3.1.1.10 Review? | <input checked="" type="checkbox"/> | <input type="checkbox"/> | <input type="checkbox"/> |
| 3.1.1.11 Proficiency testing? | <input checked="" type="checkbox"/> | <input type="checkbox"/> | <input type="checkbox"/> |
| 3.1.1.12 Corrective action? | <input checked="" type="checkbox"/> | <input type="checkbox"/> | <input type="checkbox"/> |

- | | | | |
|------------------------------|-------------------------------------|-------------------------------------|--------------------------|
| 3.1.1.13 Audits? | <input checked="" type="checkbox"/> | <input type="checkbox"/> | <input type="checkbox"/> |
| 3.1.1.14 Safety? | <input checked="" type="checkbox"/> | <input type="checkbox"/> | <input type="checkbox"/> |
| 3.1.1.15 Outsourcing? | <input type="checkbox"/> | <input checked="" type="checkbox"/> | <input type="checkbox"/> |

[Click Here For Discussion](#)

Comment

Standards 3.1.1, 3.1.1.4, 3.1.1.15 - See Findings Section

- | | | Yes | No | N/A |
|------------|---|---|-----------------------------|--------------------------|
| 3.2 | Does the laboratory maintain and follow a procedure regarding document retention that specifically addresses: | <input checked="" type="checkbox"/> | <input type="checkbox"/> | <input type="checkbox"/> |
| a. | Proficiency tests? | Yes <input checked="" type="checkbox"/> | No <input type="checkbox"/> | |
| b. | Corrective action? | Yes <input checked="" type="checkbox"/> | No <input type="checkbox"/> | |
| c. | Audits? | Yes <input checked="" type="checkbox"/> | No <input type="checkbox"/> | |
| d. | Training records? | Yes <input checked="" type="checkbox"/> | No <input type="checkbox"/> | |
| e. | Continuing education? | Yes <input checked="" type="checkbox"/> | No <input type="checkbox"/> | |
| f. | Case files? | Yes <input checked="" type="checkbox"/> | No <input type="checkbox"/> | |
| g. | Court testimony monitoring? | Yes <input checked="" type="checkbox"/> | No <input type="checkbox"/> | |

[Click Here For Discussion](#)

Comment

3.3 Is the quality system as applicable to DNA reviewed annually (calendar year) independent of the audit required by Standard 15, and is the review performed under the direction and documented approval of the technical leader?

| Yes | No | N/A |
|-------------------------------------|--------------------------|--------------------------|
| <input checked="" type="checkbox"/> | <input type="checkbox"/> | <input type="checkbox"/> |

[Click Here For Discussion](#)

Comment

Standard 4. Organization and Management

| | Yes | No | N/A |
|---|-------------------------------------|-------------------------------------|-------------------------------------|
| 4.1 Does the laboratory have: | <input type="checkbox"/> | <input checked="" type="checkbox"/> | <input type="checkbox"/> |
| 4.1.1 A managerial staff with the authority and resources needed to discharge its duties and meet the requirements of the Standards in this document? | <input checked="" type="checkbox"/> | <input type="checkbox"/> | <input type="checkbox"/> |
| 4.1.2 A technical leader who is accountable for the technical operations? | <input type="checkbox"/> | <input checked="" type="checkbox"/> | <input type="checkbox"/> |
| a. Have at least one technical leader in a multi-laboratory system? | <input type="checkbox"/> | <input type="checkbox"/> | <input checked="" type="checkbox"/> |
| 4.1.3 A casework CODIS administrator who is accountable for CODIS on-site at each individual laboratory facility using CODIS? | <input type="checkbox"/> | <input checked="" type="checkbox"/> | <input type="checkbox"/> |
| 4.1.4 At least two full-time employees who are qualified DNA analysts? | <input checked="" type="checkbox"/> | <input type="checkbox"/> | <input type="checkbox"/> |
| 4.1.5 Documentation that specifies the responsibility, authority, and interrelation of all personnel who manage, perform, or verify work affecting the validity of the DNA analysis? | <input type="checkbox"/> | <input checked="" type="checkbox"/> | <input type="checkbox"/> |
| 4.1.6 A documented contingency plan that is approved by laboratory management if the technical leader position is vacated? | <input checked="" type="checkbox"/> | <input type="checkbox"/> | <input type="checkbox"/> |

[Click Here For Discussion](#)

Comment

Standards 4.1, 4.1.2, 4.1.3 and 4.1.5 - See Findings Section

Standard 4.1.2.a was rated N/A as the laboratory is not part of a multi-laboratory system.

Standard 5. Personnel

- | | Yes | No | N/A |
|---|--------------------------|-------------------------------------|--------------------------|
| 5.1 Do laboratory personnel have the education, training, and experience commensurate with the examination and testimony provided? | <input type="checkbox"/> | <input checked="" type="checkbox"/> | <input type="checkbox"/> |

[Click Here For Discussion](#)

Comment

Standard 5.1 - See Finding Section

- | | Yes | No | N/A |
|---|--------------------------|-------------------------------------|--------------------------|
| 5.1.1 Does the laboratory have written job descriptions for all personnel to include responsibilities, duties, and skills? | <input type="checkbox"/> | <input checked="" type="checkbox"/> | <input type="checkbox"/> |

[Click Here For Discussion](#)

Comment

Standard 5.1.1 - See Findings Section

- | | Yes | No | N/A |
|--|-------------------------------------|-------------------------------------|--------------------------|
| 5.1.2 Does the laboratory have a documented training program for qualifying all analyst(s) and technician(s)? | <input checked="" type="checkbox"/> | <input type="checkbox"/> | <input type="checkbox"/> |
| 5.1.2.1 Does the training program contain at a minimum the following components: | | | |
| a. A training manual that covers all applicable DNA analytical procedures that the analyst/technician will perform? | <input type="checkbox"/> | <input checked="" type="checkbox"/> | <input type="checkbox"/> |
| b. Practical exercises that include the examination of a range of samples routinely encountered in casework? | <input checked="" type="checkbox"/> | <input type="checkbox"/> | <input type="checkbox"/> |

- 5.1.2.2 Does the laboratory’s training program teach and assess the technical skills and knowledge required to perform DNA analysis and include, at a minimum, the following?
- 5.1.2.2.1 Does the training program require the documentation of the successful completion of a competency test(s)?
- 5.1.2.2.2 For an analyst or technician with previous forensic experience:
 - a. Did the technical leader assess and document the adequacy of the previous training of the analyst and/or technician?
 - b. Did the analyst and/or technician complete a modified training program that was assessed and documented by the technical leader?
- 5.1.2.2.3 Prior to participating in independent casework did all analysts and technicians, regardless of previous experience, successfully complete a competency test(s) covering the routine DNA methodologies to be used?

[Click Here For Discussion](#)

Comment

Standards 5.1.2.1.a and 5.1.2.2 - See Findings Section

Standards 5.1.2.2.2.a and 5.1.2.2.2.b were rated N/A as no analysts or technicians with previous forensic experience were hired and put through a modified training program since the last external audit.

- | | | Yes | No | N/A |
|---------|--|-------------------------------------|--------------------------|--------------------------|
| 5.1.3 | Does the laboratory have a documented program to ensure that technical qualifications are maintained through continuing education? | <input checked="" type="checkbox"/> | <input type="checkbox"/> | <input type="checkbox"/> |
| 5.1.3.1 | Does the technical leader, casework CODIS administrator, and each analyst have documented attendance at seminars, courses, professional meetings, or documented training sessions/classes that consist of: | <input checked="" type="checkbox"/> | <input type="checkbox"/> | <input type="checkbox"/> |

a. Subject areas relevant to the developments in DNA typing?

Yes No

b. Cumulative minimum of eight hours per calendar year?

Yes No

5.1.3.1.1 For continuing education conducted internally, does the laboratory's retained documentation include the following:

a. Title of the program? Yes No

b. A record of the presentation? Yes No

c. Date of the training? Yes No

d. Attendance list? Yes No

e. Curriculum vitae of the presenter(s)? Yes No

5.1.3.1.2 For continuing education conducted externally, does the laboratory's retained documentation include one or more of the following:

a. Certificate of attendance?

b. Program agenda/syllabus?

c. Travel documentation?

5.1.3.1.3 For continuing education that is based on multimedia or Internet delivery:

a. Was the training subject to the review of, and approved by, the technical leader?
Yes No

b. Was the time required to complete the program formally recorded in the laboratory's retained document?
Yes No

c. Was the completion submitted to the technical leader for review and approval?
Yes No

5.1.3.2 For the review of scientific literature:

- a. Does the laboratory have a program, approved by the technical leader, for the annual review of scientific literature that documents the ongoing reading of scientific literature?
- b. Does the laboratory maintain or have physical or electronic access to a collection of current books, reviewed journals, or other literature applicable to DNA analysis?

[Click Here For Discussion](#)

Comment

Standard 5.1.3.2.a - See Findings Section

- | | Yes | No | N/A |
|--|-------------------------------------|--------------------------|--------------------------|
| 5.1.4 Does the laboratory maintain records on the relevant qualifications, training, skills, and experience of all technical personnel? | <input checked="" type="checkbox"/> | <input type="checkbox"/> | <input type="checkbox"/> |

[Click Here For Discussion](#)

Comment

- | | Yes | No | N/A |
|---|-------------------------------------|-------------------------------------|--------------------------|
| 5.2 Does the technical leader satisfy the requirements for degree/education, experience, and duties listed in Standards 5.2.1 through 5.2.4.1? | <input type="checkbox"/> | <input checked="" type="checkbox"/> | <input type="checkbox"/> |
| 5.2.1 Does the technical leader of the laboratory meet or exceed the following degree/educational requirements? | <input type="checkbox"/> | <input checked="" type="checkbox"/> | <input type="checkbox"/> |
| a. A master's degree in a biology-, chemistry-, or forensic science-related area or have a waiver as stated in Standard 5.2.1.4? | <input checked="" type="checkbox"/> | <input type="checkbox"/> | <input type="checkbox"/> |

b. Twelve semester hours or equivalent credit hours including a combination of graduate and undergraduate course work or classes covering the following subject areas:

- 1. Biochemistry? Yes No
- 2. Genetics? Yes No
- 3. Molecular biology? Yes No
- 4. Statistics or population genetics? Yes No

5.2.1.1 Of the 12 semester or equivalent credit hours required, do they include at least one graduate-level course registering 3 or more semester or equivalent credit hours?

5.2.1.2 Do each of the specific subject areas listed in Standard 5.2.1 constitute an integral component of any coursework used to demonstrate compliance with this Standard?

5.2.1.3 For individuals who have completed coursework with titles other than those listed in Standard 5.2.1, have they successfully demonstrated compliance with this Standard through a combination of pertinent materials such as a transcript, syllabus, letter from the instructor, or other documentation that supports the course content?

[Click Here For Discussion](#)

Comment

Standard 5.2 - The current DNA Technical Leader meets all criteria listed above. The NO ratings are referring to the mtDNA Technical Leader that was in place from January - August 2009.

Standards 5.2, 5.2.1, 5.2.1.b, 5.2.1.b.4, and 5.2.1.1 - See Findings Section

- | | Yes | No | N/A |
|--|--------------------------|--------------------------|-------------------------------------|
| 5.2.1.4 If the degree requirements of Standard 5.2.1 are not met, does the technical leader possess a waiver from the American Society of Crime Laboratory Directors (ASCLD)? | <input type="checkbox"/> | <input type="checkbox"/> | <input checked="" type="checkbox"/> |

[Click Here For Discussion](#)

Comment

Standard 5.2.1.4 was rated N/A since the current Technical Leader does not possess a waiver from ASCLD; the current Technical Leader meets the degree requirements of Standard 5.2.1.

See standard 5.2 for mtDNA Technical Leader (Jan-Aug 2009) degree requirements.

- | | Yes | No | N/A |
|---|-------------------------------------|--------------------------|--------------------------|
| 5.2.2 Technical leader minimum experience requirements: | | | |
| a. Does the technical leader have three years of forensic DNA laboratory experience obtained at a laboratory where forensic DNA testing was conducted for the identification and evaluation of biological evidence in criminal matters? | <input checked="" type="checkbox"/> | <input type="checkbox"/> | <input type="checkbox"/> |
| b. Does any technical leader, appointed or hired on or after July 1, 2009, have a minimum of three years human-DNA experience (current or previous) as a qualified analyst on forensic samples? | <input checked="" type="checkbox"/> | <input type="checkbox"/> | <input type="checkbox"/> |
| c. Has the technical leader successfully completed, or will successfully complete within one year of appointment, the FBI-sponsored auditor training? | <input checked="" type="checkbox"/> | <input type="checkbox"/> | <input type="checkbox"/> |

[Click Here For Discussion](#)

Comment

The above ratings are for the current Technical Leader.

| | Yes | No | N/A |
|---|-------------------------------------|-------------------------------------|--------------------------|
| 5.2.3 Does the technical leader of the laboratory have responsibility for the following: | | | |
| 5.2.3.1 Does the technical leader have the following general duties and authority: | | | |
| 5.2.3.1.1 Oversee the technical operations of the laboratory? | <input checked="" type="checkbox"/> | <input type="checkbox"/> | <input type="checkbox"/> |
| 5.2.3.1.2 Authority to initiate, suspend, and resume DNA analytical operations for the laboratory or an individual? | <input type="checkbox"/> | <input checked="" type="checkbox"/> | <input type="checkbox"/> |
| 5.2.3.2 Does the technical leader perform the following specific responsibilities: | | | |
| 5.2.3.2.1 Evaluate and document approval of all validations and methods used by the laboratory and propose new or modified analytical procedures to be used by analysts? | <input type="checkbox"/> | <input checked="" type="checkbox"/> | <input type="checkbox"/> |
| 5.2.3.2.2 Review and document the review of the academic transcripts and training records for newly qualified analysts and approve their qualifications prior to their conducting independent casework analysis? | <input type="checkbox"/> | <input checked="" type="checkbox"/> | <input type="checkbox"/> |
| 5.2.3.2.3 Approve the technical specifications for outsourcing agreements? | <input type="checkbox"/> | <input checked="" type="checkbox"/> | <input type="checkbox"/> |
| 5.2.3.2.4 Review and document the review of internal and external DNA audit documents and, if applicable, approve corrective action(s). | <input type="checkbox"/> | <input checked="" type="checkbox"/> | <input type="checkbox"/> |
| 5.2.3.2.5 Review annually the procedures of the laboratory and document such review? | <input type="checkbox"/> | <input checked="" type="checkbox"/> | <input type="checkbox"/> |
| 5.2.3.2.6 Review and approve the training, quality assurance, and proficiency testing programs in the laboratory? | <input checked="" type="checkbox"/> | <input type="checkbox"/> | <input type="checkbox"/> |

Comment

[Click Here For Discussion](#)

Standards 5.2.3.1.2, 5.2.3.2.1, 5.2.3.2.2, 5.2.3.2.3, 5.2.3.2.4 and 5.2.3.2.5 - See Findings Section

The above ratings are for the current Technical Leader.

| | Yes | No | N/A |
|---|-------------------------------------|-------------------------------------|-------------------------------------|
| 5.2.4 Technical leader accessibility: | | | |
| a. Is the technical leader accessible to the laboratory to provide on-site, telephonic, or electronic consultation as needed? | <input checked="" type="checkbox"/> | <input type="checkbox"/> | <input type="checkbox"/> |
| b. If the technical leader oversees a system of separate laboratories, has the technical leader conducted semiannual on-site visits of each of the laboratories? | <input type="checkbox"/> | <input type="checkbox"/> | <input checked="" type="checkbox"/> |
| 5.2.4.1 Is the technical leader a full-time employee of the laboratory or laboratory system? | <input checked="" type="checkbox"/> | <input type="checkbox"/> | <input type="checkbox"/> |
| 5.2.4.1.1 a. If the technical leader position of the laboratory had been vacant since the last audit, was there a qualified individual immediately appointed as technical leader? | <input type="checkbox"/> | <input type="checkbox"/> | <input checked="" type="checkbox"/> |
| b. If a qualified individual was not available/ appointed, did the laboratory immediately contact the FBI and submit its contingency plan within 14 days of the vacancy for the FBI's approval? | <input type="checkbox"/> | <input type="checkbox"/> | <input checked="" type="checkbox"/> |
| c. Was all new casework suspended until the plan was approved by the FBI? | <input type="checkbox"/> | <input type="checkbox"/> | <input checked="" type="checkbox"/> |
| 5.2.5 Did each technical leader appointed or hired on or after July 1, 2009, document a review of the following: | | | |
| 5.2.5.1 Validation studies and methodologies currently used by the laboratory? | <input type="checkbox"/> | <input checked="" type="checkbox"/> | <input type="checkbox"/> |
| 5.2.5.2 Educational qualifications and training records of currently qualified analysts? | <input checked="" type="checkbox"/> | <input type="checkbox"/> | <input type="checkbox"/> |

[Click Here For Discussion](#)

Comment

Standard 5.2.5.1 - See Findings Section

Standard 5.2.4.b was rated N/A as the Technical Leader does not oversee a system of separate laboratories.

Standards 5.2.4.1.1.a, 5.2.4.1.1.b, and 5.2.4.1.1.c were rated N/A as the Technical Leader position has not been vacant since the last external audit.

The above ratings are for the current Technical Leader.

| | | Yes | No | N/A |
|--------------|--|-------------------------------------|--------------------------|-------------------------------------|
| 5.3 | Is the casework CODIS administrator an employee of the laboratory and does he or she meet the following qualifications? | <input checked="" type="checkbox"/> | <input type="checkbox"/> | <input type="checkbox"/> |
| 5.3.1 | Education: | | | |
| | Does the casework CODIS administrator meet the minimum education requirements? | <input checked="" type="checkbox"/> | <input type="checkbox"/> | <input type="checkbox"/> |
| | a. Does the casework CODIS administrator meet the minimum education requirements as defined in Standard 5.4 or | <input checked="" type="checkbox"/> | <input type="checkbox"/> | <input type="checkbox"/> |
| | b. Was the casework CODIS administrator appointed or hired prior to July 1, 2009, with supporting documentation from the FBI? | <input type="checkbox"/> | <input type="checkbox"/> | <input checked="" type="checkbox"/> |
| 5.3.2 | Experience: | | | |
| | Does the casework CODIS administrator meet the experience requirements? | <input checked="" type="checkbox"/> | <input type="checkbox"/> | <input type="checkbox"/> |
| | a. Is a current or previously qualified casework DNA analyst with documented mixture interpretation training, or | <input checked="" type="checkbox"/> | <input type="checkbox"/> | <input type="checkbox"/> |
| | b. Was the casework CODIS administrator appointed or hired prior to July 1, 2009 with documented mixture-interpretation training and completion of FBI-sponsored CODIS training? | <input type="checkbox"/> | <input type="checkbox"/> | <input checked="" type="checkbox"/> |

[Click Here For Discussion](#)

Comment

Standards 5.3.1.b and 5.3.2.b were rated as N/A as the casework CODIS administrator meets the minimum education requirements as defined in Standard 5.4 and has documented mixture interpretation training.

| | Yes | No | N/A |
|--|-------------------------------------|-------------------------------------|-------------------------------------|
| 5.3.3 Has the casework CODIS administrator: | | | |
| a. Successfully completed the FBI auditor training within one year of appointment, if not previously attended such training? | <input checked="" type="checkbox"/> | <input type="checkbox"/> | <input type="checkbox"/> |
| b. Participated in the FBI sponsored CODIS software training within six months of appointment, if not previously attended such training? | <input checked="" type="checkbox"/> | <input type="checkbox"/> | <input type="checkbox"/> |
| 5.3.4 Is the casework CODIS administrator responsible for the following: | <input type="checkbox"/> | <input checked="" type="checkbox"/> | <input type="checkbox"/> |
| 5.3.4.1 Administering the laboratory's local CODIS network? | <input checked="" type="checkbox"/> | <input type="checkbox"/> | <input type="checkbox"/> |
| 5.3.4.2 Scheduling and documenting the CODIS computer training of casework analysts? | <input type="checkbox"/> | <input checked="" type="checkbox"/> | <input type="checkbox"/> |
| 5.3.4.3 Assuring that the security of data stored in CODIS is in accordance with state and/or federal law and NDIS operational procedures? | <input type="checkbox"/> | <input checked="" type="checkbox"/> | <input type="checkbox"/> |
| 5.3.4.4 Assuring that the quality of data stored in CODIS is in accordance with state and/or federal law and NDIS operational procedures? | <input type="checkbox"/> | <input checked="" type="checkbox"/> | <input type="checkbox"/> |
| 5.3.4.5 Assuring that matches are dispositioned in accordance with NDIS operational procedures? | <input type="checkbox"/> | <input checked="" type="checkbox"/> | <input type="checkbox"/> |
| 5.3.5 Is the casework CODIS administrator authorized to terminate an analyst's or the laboratory's participation in CODIS until the reliability and security of the computer data can be assured if an issue with the data is identified? | <input type="checkbox"/> | <input checked="" type="checkbox"/> | <input type="checkbox"/> |
| 5.3.6 If the casework CODIS administrator position has been unoccupied since the last audit, has the laboratory refrained from uploading new DNA profiles to NDIS during the vacancy? | <input type="checkbox"/> | <input type="checkbox"/> | <input checked="" type="checkbox"/> |

[Click Here For Discussion](#)

Comment

Standards 5.3.4, 5.3.4.2, 5.3.4.3, 5.3.4.4, 5.3.4.5 and 5.3.5 - See Findings Section

Standard 5.3.6 was rated N/A as the casework CODIS Administrator position has not been unoccupied since the last external audit.

| | | Yes | No | N/A |
|----------------|---|-------------------------------------|--------------------------|--------------------------|
| 5.4.2 | Does each analyst have six months of documented, forensic human-DNA laboratory experience? | <input checked="" type="checkbox"/> | <input type="checkbox"/> | <input type="checkbox"/> |
| 5.4.2.1 | Prior to independent work using DNA technology, has each analyst completed the analysis of a range of samples routinely encountered in forensic casework? | <input checked="" type="checkbox"/> | <input type="checkbox"/> | <input type="checkbox"/> |
| 5.4.2.2 | Has each analyst successfully completed a competency test before beginning independent DNA analysis? | <input checked="" type="checkbox"/> | <input type="checkbox"/> | <input type="checkbox"/> |

[Click Here For Discussion](#)

Comment

| | | Yes | No | N/A |
|--------------|---|-------------------------------------|--------------------------|-------------------------------------|
| 5.5 | Has each technician successfully completed each of the following: | | | |
| 5.5.1 | Documented training specific to his or her job function(s)? | <input checked="" type="checkbox"/> | <input type="checkbox"/> | <input type="checkbox"/> |
| 5.5.2 | A competency test before participating in DNA analysis on evidence? | <input checked="" type="checkbox"/> | <input type="checkbox"/> | <input type="checkbox"/> |
| 5.6 | Do all laboratory technical support personnel have documented training specific to their job function(s)? | <input type="checkbox"/> | <input type="checkbox"/> | <input checked="" type="checkbox"/> |

[Click Here For Discussion](#)

Comment

Standard 5.6 was rated N/A as the laboratory does not utilize technical support personnel for DNA analysis.

Standard 6. Facilities

| | Yes | No | N/A |
|--|-------------------------------------|--------------------------|--------------------------|
| 6.1 Is the laboratory designed to ensure the integrity of the analyses and the evidence? | <input checked="" type="checkbox"/> | <input type="checkbox"/> | <input type="checkbox"/> |
| 6.1.1 Is access to the laboratory controlled and limited in a manner that prevents access by unauthorized personnel? | <input checked="" type="checkbox"/> | <input type="checkbox"/> | <input type="checkbox"/> |
| a. Do all exterior entrance/exit points have security control? | <input checked="" type="checkbox"/> | <input type="checkbox"/> | <input type="checkbox"/> |
| b. Is the distribution of all keys, combinations, and other security devices, documented and limited to the personnel designated by laboratory management? | <input checked="" type="checkbox"/> | <input type="checkbox"/> | <input type="checkbox"/> |

[Click Here For Discussion](#)

Comment

| | Yes | No | N/A |
|---|-------------------------------------|--------------------------|--------------------------|
| 6.1.2 Except as provided in Standard 6.1.4, are techniques performed prior to polymerase chain reaction (PCR) amplification-- to include evidence examinations, DNA extractions, and PCR setup-- conducted at separate times or in separate spaces from one another? | <input checked="" type="checkbox"/> | <input type="checkbox"/> | <input type="checkbox"/> |
| 6.1.3 Except as provided in Standard 6.1.4, is amplified DNA product-- including real-time PCR-- generated, processed, and maintained in a room(s) separate from the evidence examination, DNA extractions, and PCR-setup areas? | <input checked="" type="checkbox"/> | <input type="checkbox"/> | <input type="checkbox"/> |
| a. Are the doors between rooms containing amplified DNA and other areas closed at all times except for passage? | <input checked="" type="checkbox"/> | <input type="checkbox"/> | <input type="checkbox"/> |

6.1.4 If a robotic workstation is used to carry out DNA extraction, quantification, PCR setup, and/or amplification in a single room, has the laboratory validated the analytical process in accordance with Standard 8?

a. If the robot performs analysis through amplification, is the robot housed in a separate room from that used for initial evidence examinations?

[Click Here For Discussion](#)

Comment

Standards 6.1.4 and 6.1.4.a were marked N/A as the laboratory only uses robotics systems for DNA extractions.

6.1.5 Does the laboratory have and follow written procedures for cleaning and decontaminating facilities and equipment? **Yes** **No** **N/A**

[Click Here For Discussion](#)

Comment

STANDARD 7 Evidence

| | | Yes | No | N/A |
|-------|--|-------------------------------------|-------------------------------------|--------------------------|
| 7.1 | Does the laboratory have and follow a documented evidence control system to ensure the integrity of physical evidence? | <input type="checkbox"/> | <input checked="" type="checkbox"/> | <input type="checkbox"/> |
| 7.1.1 | For evidence and sample identification: | | | |
| a. | Is all evidence marked with a unique identifier on the evidence package? | <input checked="" type="checkbox"/> | <input type="checkbox"/> | <input type="checkbox"/> |
| b. | Does the laboratory clearly define what constitutes evidence and what constitutes work product? | <input type="checkbox"/> | <input checked="" type="checkbox"/> | <input type="checkbox"/> |
| c. | Does the laboratory have and follow a method to distinguish each sample throughout processing? | <input checked="" type="checkbox"/> | <input type="checkbox"/> | <input type="checkbox"/> |

[Click Here For Discussion](#)

Comment

Standards 7.1 and 7.1.1.b - See Findings Section

| | | Yes | No | N/A |
|-------|---|-------------------------------------|--------------------------|--------------------------|
| 7.1.2 | Does the laboratory document and maintain a chain of custody, in hard or electronic format, for all evidence, to include the following: | <input checked="" type="checkbox"/> | <input type="checkbox"/> | <input type="checkbox"/> |
| a. | Signature or initials or the electronic equivalent of each individual receiving or transferring the evidence? | | | |
| | Yes <input checked="" type="checkbox"/> No <input type="checkbox"/> | | | |
| b. | The corresponding date for each transfer? | | | |
| | Yes <input checked="" type="checkbox"/> No <input type="checkbox"/> | | | |
| c. | Evidentiary item(s) transferred? | | | |
| | Yes <input checked="" type="checkbox"/> No <input type="checkbox"/> | | | |

[Click Here For Discussion](#)

Comment

- | | Yes | No | N/A |
|--|-------------------------------------|--------------------------|--------------------------|
| 7.1.3 Does the laboratory have and follow documented procedures designed to minimize loss, contamination, and/or deleterious change of evidence and work product in progress? | <input checked="" type="checkbox"/> | <input type="checkbox"/> | <input type="checkbox"/> |
| 7.1.4 Does the laboratory have secure, controlled-access areas for evidence storage and work product in progress? | <input checked="" type="checkbox"/> | <input type="checkbox"/> | <input type="checkbox"/> |

[Click Here For Discussion](#)

Comment

- | | Yes | No | N/A |
|---|-------------------------------------|--------------------------|--------------------------|
| 7.2 Does the laboratory retain or return a portion of the evidence sample or extract where possible? | <input checked="" type="checkbox"/> | <input type="checkbox"/> | <input type="checkbox"/> |

[Click Here For Discussion](#)

Comment

- | | Yes | No | N/A |
|--|--------------------------|-------------------------------------|--------------------------|
| 7.3 Does the laboratory have and follow documented policies for the disposition of evidence and sample consumption? | <input type="checkbox"/> | <input checked="" type="checkbox"/> | <input type="checkbox"/> |

[Click Here For Discussion](#)

Comment

Standard 7.3 - See Findings Section

Standard 8. Validation

- | | Yes | No | N/A |
|--|--------------------------|-------------------------------------|--------------------------|
| 8.1 Does the laboratory use validated methods for DNA analyses? | <input type="checkbox"/> | <input checked="" type="checkbox"/> | <input type="checkbox"/> |

[Click Here For Discussion](#)

Comment

Standard 8.1 - See Findings Section

- | | Yes | No | N/A |
|---|-------------------------------------|--------------------------|--------------------------|
| 8.2 Have developmental validation studies preceded the use of a novel methodology for forensic DNA analysis? | <input checked="" type="checkbox"/> | <input type="checkbox"/> | <input type="checkbox"/> |

[Click Here For Discussion](#)

Comment

| | | Yes | No | N/A |
|--------------|---|---|-----------------------------|------------------------------|
| 8.2.1 | Have developmental validation studies been performed and documented to include, where applicable: | <input checked="" type="checkbox"/> | <input type="checkbox"/> | <input type="checkbox"/> |
| a. | Characterization of the genetic marker? | Yes <input checked="" type="checkbox"/> | No <input type="checkbox"/> | N/A <input type="checkbox"/> |
| b. | Species specificity? | Yes <input checked="" type="checkbox"/> | No <input type="checkbox"/> | N/A <input type="checkbox"/> |
| c. | Sensitivity studies? | Yes <input checked="" type="checkbox"/> | No <input type="checkbox"/> | N/A <input type="checkbox"/> |
| d. | Stability studies? | Yes <input checked="" type="checkbox"/> | No <input type="checkbox"/> | N/A <input type="checkbox"/> |
| e. | Reproducibility? | Yes <input checked="" type="checkbox"/> | No <input type="checkbox"/> | N/A <input type="checkbox"/> |
| f. | Case-type samples? | Yes <input checked="" type="checkbox"/> | No <input type="checkbox"/> | N/A <input type="checkbox"/> |
| g. | Population studies? | Yes <input checked="" type="checkbox"/> | No <input type="checkbox"/> | N/A <input type="checkbox"/> |
| h. | Mixture studies? | Yes <input checked="" type="checkbox"/> | No <input type="checkbox"/> | N/A <input type="checkbox"/> |
| i. | Precision and accuracy studies? | Yes <input checked="" type="checkbox"/> | No <input type="checkbox"/> | N/A <input type="checkbox"/> |
| j. | PCR-based studies to include? | Yes <input checked="" type="checkbox"/> | No <input type="checkbox"/> | N/A <input type="checkbox"/> |
| | 1. Reaction conditions? | Yes <input checked="" type="checkbox"/> | No <input type="checkbox"/> | |
| | 2. Assessment of differential and preferential amplification? | Yes <input checked="" type="checkbox"/> | No <input type="checkbox"/> | |
| | 3. Effects of multiplexing? | Yes <input checked="" type="checkbox"/> | No <input type="checkbox"/> | |
| | 4. Assessment of appropriate controls? | Yes <input checked="" type="checkbox"/> | No <input type="checkbox"/> | |
| | 5. Product detection studies? | Yes <input checked="" type="checkbox"/> | No <input type="checkbox"/> | |
| 8.2.2 | Are peer-reviewed publication(s) of the underlying scientific principle(s) of a technology available? | <input checked="" type="checkbox"/> | <input type="checkbox"/> | <input type="checkbox"/> |

[Click Here For Discussion](#)

Comment

| | | Yes | No | N/A |
|----------------|---|-------------------------------------|-------------------------------------|-------------------------------------|
| 8.3 | Except as provided in Standard 8.3.1.1, have internal validation of all manual and robotic methodologies been conducted by each laboratory and reviewed and approved by the laboratory's technical leader prior to use? | <input checked="" type="checkbox"/> | <input type="checkbox"/> | <input type="checkbox"/> |
| 8.3.1 | For Internal Validation Studies: | | | |
| | a. Have internal validation studies been documented and summarized? | <input checked="" type="checkbox"/> | <input type="checkbox"/> | <input type="checkbox"/> |
| | b. Have all internal validation studies conducted on or after July 1, 2009, included, as applicable: | <input type="checkbox"/> | <input checked="" type="checkbox"/> | <input type="checkbox"/> |
| | 1. Known and non probative evidence samples or mock evidence samples? | | | |
| | Yes <input type="checkbox"/> No <input checked="" type="checkbox"/> N/A <input type="checkbox"/> | | | |
| | 2. Reproducibility and precision? | | | |
| | Yes <input checked="" type="checkbox"/> No <input type="checkbox"/> N/A <input type="checkbox"/> | | | |
| | 3. Sensitivity and stochastic studies? | | | |
| | Yes <input checked="" type="checkbox"/> No <input type="checkbox"/> N/A <input type="checkbox"/> | | | |
| | 4. Mixture studies? | | | |
| | Yes <input checked="" type="checkbox"/> No <input type="checkbox"/> N/A <input type="checkbox"/> | | | |
| | 5. Contamination assessment? | | | |
| | Yes <input type="checkbox"/> No <input checked="" type="checkbox"/> N/A <input type="checkbox"/> | | | |
| 8.3.1.1 | For multilaboratory systems: | | | |
| | a. Has each laboratory in a multi-laboratory system completed, documented, and maintained applicable site-specific precision, sensitivity, and contamination assessment studies? | <input type="checkbox"/> | <input type="checkbox"/> | <input checked="" type="checkbox"/> |
| | b. Are the summaries of all applicable validation data available at each site? | <input type="checkbox"/> | <input type="checkbox"/> | <input checked="" type="checkbox"/> |
| 8.3.2 | Have quality assurance parameters and interpretation guidelines, including, as applicable, guidelines for mixture interpretation, been defined pursuant to internal validation? | <input checked="" type="checkbox"/> | <input type="checkbox"/> | <input type="checkbox"/> |
| 8.3.3 | If a laboratory has had a change in detection platform or test kit, have internal validation studies been performed? | <input checked="" type="checkbox"/> | <input type="checkbox"/> | <input type="checkbox"/> |
| 8.4 | Has the analyst or examination team successfully completed a competency test using the DNA analysis procedure prior to its incorporation into casework applications? | <input checked="" type="checkbox"/> | <input type="checkbox"/> | <input type="checkbox"/> |

[Click Here For Discussion](#)

Comment

Standards 8.3.1.b, 8.3.1.b.1 and 8.3.1.b.5 - See Findings Section

Standards 8.3.1.1.a and 8.3.1.1.b were rated N/A as the laboratory is not part of a multi-laboratory system.

| | Yes | No | N/A |
|--|--------------------------|-------------------------------------|-------------------------------------|
| 8.5 Have modified procedures been evaluated by comparison with the original procedures using similar DNA samples prior to their incorporation into casework applications? | <input type="checkbox"/> | <input checked="" type="checkbox"/> | <input type="checkbox"/> |
| 8.6 Has the laboratory evaluated each additional or modified critical instrument by conducting a performance check prior to its use in casework? | <input type="checkbox"/> | <input checked="" type="checkbox"/> | <input type="checkbox"/> |
| 8.7 Has the laboratory evaluated software upgrades by conducting a performance check prior to use in casework? | <input type="checkbox"/> | <input type="checkbox"/> | <input checked="" type="checkbox"/> |
| a. Has new software or significant software modifications been documented and subjected to validation testing prior to use in casework? | <input type="checkbox"/> | <input type="checkbox"/> | <input checked="" type="checkbox"/> |

[Click Here For Discussion](#)

Comment

Standards 8.5 and 8.6 - See Findings Section

Standards 8.7 and 8.7.a were rated N/A as there were no upgrades, modifications or new software implemented in the laboratory since the last external audit.

Standard 9. Analytical Procedures

| | Yes | No | N/A |
|---|-------------------------------------|-------------------------------------|--------------------------|
| 9.1 Does the laboratory have and follow written analytical procedures approved by the technical leader? | <input type="checkbox"/> | <input checked="" type="checkbox"/> | <input type="checkbox"/> |
| a. Are the laboratory's standard operating procedures reviewed annually by the technical leader, and is this review documented? | <input type="checkbox"/> | <input checked="" type="checkbox"/> | <input type="checkbox"/> |
| 9.1.1 Does the laboratory have a documented standard operating procedure for each analytical method used? | <input type="checkbox"/> | <input checked="" type="checkbox"/> | <input type="checkbox"/> |
| a. Do the analytical procedures specify reagents, sample preparation, extraction methods, equipment, and controls that are standard for DNA analysis and data interpretation? | <input type="checkbox"/> | <input checked="" type="checkbox"/> | <input type="checkbox"/> |
| b. Does the laboratory have a procedure for the differential extraction of stains that contain sperm? | <input checked="" type="checkbox"/> | <input type="checkbox"/> | <input type="checkbox"/> |

[Click Here For Discussion](#)

Comment

Standards 9.1, 9.1.a, 9.1.1 and 9.1.1.a - See Findings Section

| | Yes | No | N/A |
|---|---|--|--------------------------|
| 9.2 Does the laboratory use reagents that are suitable for the methods employed? | <input type="checkbox"/> | <input checked="" type="checkbox"/> | <input type="checkbox"/> |
| 9.2.1 Does the laboratory have written procedures for documenting commercial reagents and for the formulation of in-house reagents? | <input checked="" type="checkbox"/> | <input type="checkbox"/> | <input type="checkbox"/> |
| 9.2.2 Are commercial reagents labeled with: | <input type="checkbox"/> | <input checked="" type="checkbox"/> | <input type="checkbox"/> |
| a. The identity of the reagent? | Yes <input checked="" type="checkbox"/> | No <input type="checkbox"/> | |
| b. The expiration date as provided by the manufacturer or as determined by the laboratory? | Yes <input type="checkbox"/> | No <input checked="" type="checkbox"/> | |
| 9.2.3 Are in-house reagents labeled with: | <input checked="" type="checkbox"/> | <input type="checkbox"/> | <input type="checkbox"/> |

a. The identity of the reagent?
 Yes No

b. The date of the preparation or expiration or both?
 Yes No

c. The identity of the individual preparing the reagent?
 Yes No

9.3 Critical reagents shall include, but are not limited to, the reagents listed in Standards 9.3.1 and 9.3.2.

a. Has the laboratory identified critical reagents?

b. Has the laboratory evaluated critical reagents prior to use in casework?

9.3.1 Has the laboratory identified and evaluated the following:

a. Test kits or systems for performing quantitative PCR?
 Yes No N/A

b. Test kits or systems for performing genetic typing?
 Yes No N/A

9.3.2 Has the laboratory identified and evaluated the following:

a. Thermostable DNA polymerase (if not tested as test kit components under Standard 9.3.1)?
 Yes No N/A

b. Primer sets (if not tested as test kit components under Standard 9.3.1)?
 Yes No N/A

c. Allelic ladders used for genetic analysis (if not tested as test-kit components under Standard 9.3.1)?
 Yes No N/A

[Click Here For Discussion](#)

Comment

Standards 9.2, 9.2.2, 9.2.2.b, 9.3.a, 9.3.b, 9.3.1, and 9.3.1.a - See Findings Section

Standards 9.3.2.b, and 9.3.2.c were rated N/A as the laboratory tests these items as part of the test-kit components.

- | | Yes | No | N/A |
|---|-------------------------------------|--------------------------|--------------------------|
| 9.4 Does the laboratory quantify the amount of human DNA in forensic samples prior to nuclear DNA amplification? | <input checked="" type="checkbox"/> | <input type="checkbox"/> | <input type="checkbox"/> |

[Click Here For Discussion](#)

Comment

- | | Yes | No | N/A |
|--|-------------------------------------|-------------------------------------|--------------------------|
| 9.5 Does the laboratory monitor the analytical procedures using appropriate controls and standards? | <input type="checkbox"/> | <input checked="" type="checkbox"/> | <input type="checkbox"/> |
| 9.5.1 Are standards used during quantification procedures? | <input checked="" type="checkbox"/> | <input type="checkbox"/> | <input type="checkbox"/> |
| 9.5.2 For positive and negative amplification controls: | | | |
| a. Are the positive and negative amplification controls associated with the forensic samples being typed amplified concurrently with the samples at all loci using the same primers as the forensic samples? | <input checked="" type="checkbox"/> | <input type="checkbox"/> | <input type="checkbox"/> |
| b. Are the positive and negative amplification controls associated with the forensic samples being typed? | <input checked="" type="checkbox"/> | <input type="checkbox"/> | <input type="checkbox"/> |
| 9.5.3 Are reagent blank controls associated with each extraction set being analyzed as follows: | | | |
| 9.5.3.1 Extracted concurrently? | <input type="checkbox"/> | <input checked="" type="checkbox"/> | <input type="checkbox"/> |
| 9.5.3.2 Are the reagent blanks amplified using: | <input checked="" type="checkbox"/> | <input type="checkbox"/> | <input type="checkbox"/> |
| a. The same primers as the forensic sample(s)? | | | |
| Yes <input checked="" type="checkbox"/> No <input type="checkbox"/> | | | |
| b. The same instrument model as the forensic sample(s)? | | | |
| Yes <input checked="" type="checkbox"/> No <input type="checkbox"/> | | | |
| c. The same concentration conditions as required by the forensic sample(s) containing the least amount of DNA? | | | |
| Yes <input checked="" type="checkbox"/> No <input type="checkbox"/> | | | |
| 9.5.3.3 Are the reagent blanks typed using: | <input checked="" type="checkbox"/> | <input type="checkbox"/> | <input type="checkbox"/> |

- a. The same instrument model as the forensic sample(s)?
 Yes No
- b. The same injection conditions as the forensic sample(s)?
 Yes No
- c. The most sensitive volume conditions of the forensic extraction set?
 Yes No

9.5.4 Does the laboratory use allelic ladders and internal size markers for VNTR sequence PCR- based systems?

[Click Here For Discussion](#)

Comment

Standards 9.5 and 9.5.3.1 - See Findings Section

9.5.5 Does the laboratory check its DNA procedures either annually or whenever substantial changes are made to a procedure against an appropriate and available NIST standard reference material (SRM) or standard traceable to a NIST standard? Yes No N/A

[Click Here For Discussion](#)

Comment

Standard 9.5.5 - See Findings Section

| | | Yes | No | N/A |
|--------------|--|-------------------------------------|--|--------------------------|
| 9.6 | Does the laboratory have and follow written guidelines for the interpretation of data? | <input type="checkbox"/> | <input checked="" type="checkbox"/> | <input type="checkbox"/> |
| 9.6.1 | Does the laboratory verify that all control results meet the laboratory's interpretation guidelines for all reported results? | <input checked="" type="checkbox"/> | <input type="checkbox"/> | <input type="checkbox"/> |
| 9.6.2 | Has the 1996 National Research Council report and/or a court-directed method been used for the statistical interpretation of a DNA profile for a given population and/or hypothesis or relatedness, and are these calculations derived from an established population database(s) appropriate for the calculation? | <input checked="" type="checkbox"/> | <input type="checkbox"/> | <input type="checkbox"/> |
| 9.6.3 | Does the laboratory have and follow specific documented statistical interpretation guidelines if genetic analyses that are not addressed by Standard 9.6.2 are being performed? | <input checked="" type="checkbox"/> | <input type="checkbox"/> | <input type="checkbox"/> |
| 9.6.4 | Does the laboratory have and follow documented procedures for mixture interpretation to include the following: | <input type="checkbox"/> | <input checked="" type="checkbox"/> | <input type="checkbox"/> |
| | a. Major and minor contributors? | Yes <input type="checkbox"/> | No <input checked="" type="checkbox"/> | |
| | b. Inclusions and exclusions? | Yes <input type="checkbox"/> | No <input checked="" type="checkbox"/> | |
| | c. Policies for reporting results and statistics? | Yes <input type="checkbox"/> | No <input checked="" type="checkbox"/> | |

[Click Here For Discussion](#)

Comment

Standards 9.6, 9.6.4 and 9.6.4.a, 9.6.4.b, and 9.6.4.c - See Findings Section

- | | Yes | No | N/A |
|---|--------------------------|-------------------------------------|--------------------------|
| 9.7 Does the laboratory have and follow a documented policy for detecting and controlling contamination? | <input type="checkbox"/> | <input checked="" type="checkbox"/> | <input type="checkbox"/> |

[Click Here For Discussion](#)

Comment

Standard 9.7 - See Findings Section

Standard 10. Equipment Calibration and Maintenance

- | | Yes | No | N/A |
|---|-------------------------------------|-------------------------------------|--------------------------|
| 10.1 Does the laboratory use equipment that is suitable for the methods employed? | <input checked="" type="checkbox"/> | <input type="checkbox"/> | <input type="checkbox"/> |
| 10.2 Does the laboratory have and follow a documented program for conducting performance checks and calibrating equipment and instruments? | <input type="checkbox"/> | <input checked="" type="checkbox"/> | <input type="checkbox"/> |
| 10.2.1 At a minimum, are the following critical instruments or equipment performance-checked at least annually: | <input type="checkbox"/> | <input checked="" type="checkbox"/> | <input type="checkbox"/> |
| 10.2.1.1 A thermometer that is traceable to national or international standard(s) and is used for conducting performance checks? | <input checked="" type="checkbox"/> | <input type="checkbox"/> | <input type="checkbox"/> |
| 10.2.1.2 Balance/scale? | <input checked="" type="checkbox"/> | <input type="checkbox"/> | <input type="checkbox"/> |
| 10.2.1.3 Thermal cycler temperature-verification system? | <input type="checkbox"/> | <input checked="" type="checkbox"/> | <input type="checkbox"/> |
| 10.2.1.4 Thermal cycler including quantitative-PCR? | <input checked="" type="checkbox"/> | <input type="checkbox"/> | <input type="checkbox"/> |
| 10.2.1.5 Electrophoresis detection systems? | <input type="checkbox"/> | <input checked="" type="checkbox"/> | <input type="checkbox"/> |
| 10.2.1.6 Robotic systems? | <input type="checkbox"/> | <input checked="" type="checkbox"/> | <input type="checkbox"/> |
| 10.2.1.7 Genetic analyzers? | <input type="checkbox"/> | <input checked="" type="checkbox"/> | <input type="checkbox"/> |
| 10.2.1.8 Mechanical pipettes? | <input checked="" type="checkbox"/> | <input type="checkbox"/> | <input type="checkbox"/> |
| 10.3 Does the laboratory have a schedule and follow a documented program to ensure that instruments and equipment are maintained properly? | <input type="checkbox"/> | <input checked="" type="checkbox"/> | <input type="checkbox"/> |

- a. Has documentation been retained for maintenance, service, and/or calibration?
- 10.4** Does the laboratory performance check new critical instruments and equipment, or critical instruments and equipment that have undergone repair, service or calibration, before their use in casework analysis?
- 10.4.1** At a minimum, are the following critical instruments or equipment performance-checked following repair, service, or calibration:
 - 10.4.1.1** Electrophoresis detection systems?
 - 10.4.1.2** Robotic systems?
 - 10.4.1.3** Genetic analyzers?
 - 10.4.1.4** Thermal cycler including quantative-PCR?

Comment

[Click Here For Discussion](#)

Standards 10.2, 10.2.1, 10.2.1.3, 10.2.1.5, 10.2.1.6, 10.2.1.7, 10.3, 10.4 and 10.4.1.3 - See Findings Section.

Standard 10.4.1.1 was rated N/A. The laboratory uses Agilent 2100 Bioanalyzers; however there were no repairs, service or calibration performed on them since the last external audit.

Standard 11 Reports

| | Yes | No | N/A |
|--|-------------------------------------|-------------------------------------|--------------------------|
| 11.1 a. Does the laboratory have and follow written procedures for taking and maintaining case notes to support the conclusions drawn in laboratory reports? | <input type="checkbox"/> | <input checked="" type="checkbox"/> | <input type="checkbox"/> |
| b. Does the laboratory maintain all analytical documentation generated by analysts related to case analyses? | <input checked="" type="checkbox"/> | <input type="checkbox"/> | <input type="checkbox"/> |
| c. Does the laboratory retain, in hard copy or electronic format, sufficient documentation for each technical analysis to support the report conclusions such that another qualified individual could interpret and evaluate the data? | <input checked="" type="checkbox"/> | <input type="checkbox"/> | <input type="checkbox"/> |

[Click Here For Discussion](#)

Comment

Standard 11.1.a - See Findings Section

| | Yes | No | N/A |
|---|-------------------------------------|-------------------------------------|--------------------------|
| 11.2 Do the laboratory reports include the following elements: | <input type="checkbox"/> | <input checked="" type="checkbox"/> | <input type="checkbox"/> |
| 11.2.1 Case identifier? | <input checked="" type="checkbox"/> | <input type="checkbox"/> | <input type="checkbox"/> |
| 11.2.2 Description of evidence examined? | <input checked="" type="checkbox"/> | <input type="checkbox"/> | <input type="checkbox"/> |
| 11.2.3 Description of technology? | <input checked="" type="checkbox"/> | <input type="checkbox"/> | <input type="checkbox"/> |
| 11.2.4 Locus or amplification system? | <input checked="" type="checkbox"/> | <input type="checkbox"/> | <input type="checkbox"/> |
| 11.2.5 Results and/or conclusions? | <input checked="" type="checkbox"/> | <input type="checkbox"/> | <input type="checkbox"/> |
| 11.2.6 A quantitative or qualitative interpretative statement? | <input checked="" type="checkbox"/> | <input type="checkbox"/> | <input type="checkbox"/> |
| 11.2.7 Date issued? | <input checked="" type="checkbox"/> | <input type="checkbox"/> | <input type="checkbox"/> |
| 11.2.8 Disposition of evidence? | <input checked="" type="checkbox"/> | <input type="checkbox"/> | <input type="checkbox"/> |
| 11.2.9 Signature and title, or equivalent identification, of the person accepting responsibility for the content of the report? | <input type="checkbox"/> | <input checked="" type="checkbox"/> | <input type="checkbox"/> |

[Click Here For Discussion](#)

Comment

Standards 11.2 and 11.2.9 - See Findings Section

| | Yes | No | N/A |
|---|-------------------------------------|--------------------------|--------------------------|
| 11.3 Does the laboratory maintain the confidentiality of reports, case files, DNA records, and databases, except as otherwise provided by applicable state or federal law? | <input checked="" type="checkbox"/> | <input type="checkbox"/> | <input type="checkbox"/> |
| 11.3.1 Does the laboratory have and follow written procedures to ensure the privacy of reports, case files, DNA records, and databases? | <input checked="" type="checkbox"/> | <input type="checkbox"/> | <input type="checkbox"/> |
| 11.3.2 Does the laboratory have and follow written procedures for the release of reports, case files, DNA records, and databases in accordance with applicable state or federal law? | <input checked="" type="checkbox"/> | <input type="checkbox"/> | <input type="checkbox"/> |
| 11.3.3 Does the laboratory release personally identifiable information in accordance with applicable state and federal law? | <input checked="" type="checkbox"/> | <input type="checkbox"/> | <input type="checkbox"/> |

[Click Here For Discussion](#)

Comment

Standard 12. Review

| | Yes | No | N/A |
|---|-------------------------------------|-------------------------------------|--------------------------|
| 12.1 Does the laboratory conduct and document administrative and technical reviews of all case files and reports to ensure that conclusions and supporting data are reasonable and within the constraints of scientific knowledge? | <input type="checkbox"/> | <input checked="" type="checkbox"/> | <input type="checkbox"/> |
| 12.1.1 Are all technical reviews conducted by an individual that is, or has been, a qualified analyst in the methodology being reviewed? | <input checked="" type="checkbox"/> | <input type="checkbox"/> | <input type="checkbox"/> |

[Click Here For Discussion](#)

Comment

Standard 12.1 - See Findings Section

| | Yes | No | N/A |
|--|--------------------------|-------------------------------------|--------------------------|
| 12.2 Does the laboratory document the completion of the technical review of forensic casework, and does it include the following elements: | <input type="checkbox"/> | <input checked="" type="checkbox"/> | <input type="checkbox"/> |
| 12.2.1 A review of all case notes, worksheets, and electronic data (or printed electropherograms/images) that support the conclusions? | <input type="checkbox"/> | <input checked="" type="checkbox"/> | <input type="checkbox"/> |
| 12.2.2 A review of all DNA types to verify that they are supported by the raw or analyzed data (electropherograms or images)? | <input type="checkbox"/> | <input checked="" type="checkbox"/> | <input type="checkbox"/> |
| 12.2.3 A review of all profiles to verify correct inclusions and exclusions (if applicable) as well as a review of any inconclusive result for compliance with laboratory guidelines? | <input type="checkbox"/> | <input checked="" type="checkbox"/> | <input type="checkbox"/> |
| 12.2.4 A review of all controls, internal lane standards, and allelic ladders to verify that the expected results were obtained? | <input type="checkbox"/> | <input checked="" type="checkbox"/> | <input type="checkbox"/> |
| 12.2.5 A review of statistical analysis, if applicable? | <input type="checkbox"/> | <input checked="" type="checkbox"/> | <input type="checkbox"/> |

- 12.2.6 A review of the final report to verify that the results/conclusions are supported by the data?
- a. Does the report address each tested item or its probative fraction?
- 12.2.7 For verification of CODIS eligibility. Has there been verification that all profiles entered into CODIS are eligible and have the correct DNA types and correct specimen category?
- 12.2.7.1 Prior to upload to or search of SDIS, have the following been verified for DNA profiles:
- a. Eligibility for CODIS? Yes No
- b. Correct DNA types? Yes No
- c. Appropriate specimen category? Yes No
- 12.2.7.2 Prior to entry of a DNA profile into a searchable category of SDIS, were the following criteria verified by two concordant assessments by a qualified analyst or technical reviewer?
- a. Eligibility for CODIS? Yes No
- b. Correct DNA types? Yes No
- c. Appropriate specimen category? Yes No

[Click Here For Discussion](#)

Comment

Standards 12.2, 12.2.1, 12.2.2, 12.2.3, 12.2.4, 12.2.5, 12.2.6, 12.2.6.a, 12.2.7.2, 12.2.7.2.a, 12.2.7.2.b, and 12.2.7.2.c - See Findings Section

| | Yes | No | N/A |
|---|-------------------------------------|--------------------------|--------------------------|
| 12.3 Does the administrative review include the following elements (any or all of which may be included within the technical-review process): | | | |
| 12.3.1 A review of the case file and final report for clerical errors and for the presence and accuracy of the information specified in Standard 11.2? | <input checked="" type="checkbox"/> | <input type="checkbox"/> | <input type="checkbox"/> |
| 12.3.2 A review of the chain of custody and disposition of evidence? | <input checked="" type="checkbox"/> | <input type="checkbox"/> | <input type="checkbox"/> |
| 12.3.3 A procedure to document the completion of the administrative review? | <input checked="" type="checkbox"/> | <input type="checkbox"/> | <input type="checkbox"/> |

[Click Here For Discussion](#)

Comment

| | Yes | No | N/A |
|---|-------------------------------------|-------------------------------------|--------------------------|
| 12.4 Does the laboratory document the elements of a technical and administrative review? | <input type="checkbox"/> | <input checked="" type="checkbox"/> | <input type="checkbox"/> |
| a. Are case files reviewed and documented according to the laboratory's procedures? | <input type="checkbox"/> | <input checked="" type="checkbox"/> | <input type="checkbox"/> |
| 12.5 Does the laboratory have and follow a documented procedure to address unresolved discrepant conclusions between analysts and reviewers? | <input checked="" type="checkbox"/> | <input type="checkbox"/> | <input type="checkbox"/> |
| 12.6 Does the laboratory have and follow a documented procedure for the verification and resolution of database matches? | <input checked="" type="checkbox"/> | <input type="checkbox"/> | <input type="checkbox"/> |

[Click Here For Discussion](#)

Comment

Standards 12.4 and 12.4.a - See Findings Section

12.7 Does the laboratory have and follow a program that documents the annual monitoring of the testimony of each analyst?

Yes **No** **N/A**

[Click Here For Discussion](#)

Comment

Standard 13. Proficiency Testing

13.1 Do analysts, technical reviewers, technicians, and other personnel designated by the technical leader undergo semiannual external proficiency testing in each technology performed to the full extent in which they participate in casework?

Yes **No** **N/A**

[Click Here For Discussion](#)

Comment

| | | Yes | No | N/A |
|--------|---|-------------------------------------|-------------------------------------|-------------------------------------|
| 13.1.1 | Are individuals using both manual and automated methods proficiency-tested in each, at least once per year, to the full extent in which they participate in casework? | <input checked="" type="checkbox"/> | <input type="checkbox"/> | <input type="checkbox"/> |
| 13.1.2 | Have newly qualified individuals entered the external proficiency-testing program within six months of the date of their qualification? | <input checked="" type="checkbox"/> | <input type="checkbox"/> | <input type="checkbox"/> |
| 13.1.3 | Has the laboratory defined, documented, and consistently used the date that the proficiency test is performed as the received date, assigned date, submitted date, or due date? | <input checked="" type="checkbox"/> | <input type="checkbox"/> | <input type="checkbox"/> |
| 13.1.4 | Except as provided in Standard 13.1.4.1, has each analyst been assigned and completed his or her own external proficiency test? | <input checked="" type="checkbox"/> | <input type="checkbox"/> | <input type="checkbox"/> |
| | 13.1.4.1 If a team approach is used, have all analysts, technicians, and technical reviewers been proficiency-tested according to Standard 13.1? | <input type="checkbox"/> | <input type="checkbox"/> | <input checked="" type="checkbox"/> |
| 13.1.5 | Has the typing of all CODIS core loci or CODIS core sequence ranges been attempted for each technology performed as applicable? | <input checked="" type="checkbox"/> | <input type="checkbox"/> | <input type="checkbox"/> |
| 13.1.6 | Does the laboratory maintain the following records for proficiency tests: | | | |
| | 13.1.6.1 The test-set identifier? | <input checked="" type="checkbox"/> | <input type="checkbox"/> | <input type="checkbox"/> |
| | 13.1.6.2 Identity of the analyst, and other participants, if applicable? | <input checked="" type="checkbox"/> | <input type="checkbox"/> | <input type="checkbox"/> |
| | 13.1.6.3 Date of analysis and completion? | <input checked="" type="checkbox"/> | <input type="checkbox"/> | <input type="checkbox"/> |
| | 13.1.6.4 Copies of all data and notes supporting the conclusions? | <input checked="" type="checkbox"/> | <input type="checkbox"/> | <input type="checkbox"/> |
| | 13.1.6.5 The proficiency test results? | <input checked="" type="checkbox"/> | <input type="checkbox"/> | <input type="checkbox"/> |
| | 13.1.6.6 Any discrepancies noted? | <input checked="" type="checkbox"/> | <input type="checkbox"/> | <input type="checkbox"/> |
| | 13.1.6.7 Corrective actions taken? | <input checked="" type="checkbox"/> | <input type="checkbox"/> | <input type="checkbox"/> |
| 13.1.7 | Does the laboratory include, at a minimum, the following criteria for evaluating proficiency test results: | <input type="checkbox"/> | <input checked="" type="checkbox"/> | <input type="checkbox"/> |
| | 13.1.7.1 Evaluation: | | | |
| | a. Are all reported inclusions correct? | <input type="checkbox"/> | <input checked="" type="checkbox"/> | <input type="checkbox"/> |
| | b. Are all reported exclusions correct? | <input type="checkbox"/> | <input checked="" type="checkbox"/> | <input type="checkbox"/> |

Standards 13.1.7, 13.1.7.1.a, 13.1.7.1.b, and 13.1.7.1.c - See Findings Section

Standard 13.1.4.1 was rated N/A as the laboratory does not utilize a team approach for proficiency testing.

Standards 13.1.7.2 and 13.1.7.2.1 were rated N/A as no proficiency test results have been reported as inconclusive or not interpretable since the last external audit

Standard 13.1.7.3 was rated N/A as there were no discrepancies/errors since the last external audit.

Standard 13.1.7.4.1.1 was rated N/A as there were no administrative errors or corrective actions since the last external audit.

Standard 13.1.9.a was rated N/A as there were no non-administrative discrepancies since the last external audit.

- | | Yes | No | N/A |
|--|-------------------------------------|--------------------------|--------------------------|
| 13.2 Does the laboratory use an external proficiency-test provider(s) that is in compliance with the current proficiency-testing manufacturing guidelines established by the American Society of Crime Laboratory Directors/Laboratory Accreditation Board or is in compliance with the current International Organization for Standardization? | <input checked="" type="checkbox"/> | <input type="checkbox"/> | <input type="checkbox"/> |

[Click Here For Discussion](#)

Comment

Standard 14. Corrective Action

- | | Yes | No | N/A |
|---|------------------------------|--|------------------------------|
| 14.1 For a corrective action plan: | | | |
| a. Has the laboratory established and followed a corrective action plan that addresses discrepancies detected in proficiency tests and casework analysis? | <input type="checkbox"/> | <input checked="" type="checkbox"/> | <input type="checkbox"/> |
| b. Does the corrective action plan, at a minimum, address the following: | <input type="checkbox"/> | <input checked="" type="checkbox"/> | <input type="checkbox"/> |
| 1. Define what level/type of discrepancies are applicable to this practice? | Yes <input type="checkbox"/> | No <input checked="" type="checkbox"/> | N/A <input type="checkbox"/> |
| 2. Identify (when possible) the cause of the discrepancy? | Yes <input type="checkbox"/> | No <input checked="" type="checkbox"/> | N/A <input type="checkbox"/> |
| 3. Effect of the discrepancy? | Yes <input type="checkbox"/> | No <input checked="" type="checkbox"/> | N/A <input type="checkbox"/> |
| 4. Corrective actions taken? | Yes <input type="checkbox"/> | No <input checked="" type="checkbox"/> | N/A <input type="checkbox"/> |
| 5. Preventative measures taken (where applicable) to minimize its reoccurrence? | Yes <input type="checkbox"/> | No <input checked="" type="checkbox"/> | N/A <input type="checkbox"/> |

[Click Here For Discussion](#)

Yes No N/A

6. Is documentation of all corrective actions maintained in accordance with Standard 3.2?

Yes No N/A

14.2 Prior to implementation do all corrective actions have the documented approval of the technical leader?

[Click Here For Discussion](#)

Comment

Standards 14.1.a, 14.1.b, 14.1.b.1, 14.1.b.2, 14.1.b.3, 14.1.b.4, 14.1.b.5, and 14.1.b.6 - See Finding Section

Standard 15. Audits

| | | Yes | No | N/A |
|---------------|---|-------------------------------------|-------------------------------------|--------------------------|
| 15.1 | Has the laboratory been audited annually in accordance with the FBI DNA Quality Assurance Standards? | <input type="checkbox"/> | <input checked="" type="checkbox"/> | <input type="checkbox"/> |
| | Has the laboratory maintained documentation that the auditor(s) for this inspection include: | <input checked="" type="checkbox"/> | <input type="checkbox"/> | <input type="checkbox"/> |
| 15.2 | Has an external audit been conducted at least once every two years? | <input type="checkbox"/> | <input checked="" type="checkbox"/> | <input type="checkbox"/> |
| | a. By a qualified auditor? Yes <input type="checkbox"/> No <input checked="" type="checkbox"/> | | | |
| | b. By a current or previously qualified analyst in the laboratory's current DNA technologies and platform? Yes <input type="checkbox"/> No <input checked="" type="checkbox"/> | | | |
| 15.2.1 | Has the laboratory maintained audit documentation of those individuals (i.e., casework CODIS administrator, technical leader, and analysts) that have had their education, experience, and training qualifications evaluated and approved during two external audits? | <input checked="" type="checkbox"/> | <input type="checkbox"/> | <input type="checkbox"/> |
| 15.2.2 | Has the laboratory maintained the documentation for those validations previously evaluated and approved during one external audit? | <input checked="" type="checkbox"/> | <input type="checkbox"/> | <input type="checkbox"/> |
| 15.3 | For internal audits, has the laboratory maintained documentation that the auditor(s) for this inspection include: | <input type="checkbox"/> | <input checked="" type="checkbox"/> | <input type="checkbox"/> |

- a. A qualified auditor? Yes No
- b. A current or previously qualified analyst in the laboratory's current DNA technologies and platform? Yes No
- 15.4** Have the internal and/or external audits performed pursuant to Standard 15.1 been conducted using the FBI DNA Quality Assurance Standards Audit Document in effect at that time?
- 15.5** Have internal and external DNA audit documents and, if applicable, corrective action(s) been submitted to the technical leader for review to ensure that findings, if any, were appropriately addressed?
- 15.5.1** For NDIS-participating laboratories, did the laboratory provide all external audit documentation and laboratory responses to the FBI within 30 days of the laboratory's receipt of the audit documents or report?
- 15.6** Are previous internal and external audit documents retained and available for auditor inspection?

[Click Here For Discussion](#)

Comment

Standards 15.1 (first section), 15.2, 15.2.a, 15.2.b, 15.3, and 15.3.b - See Findings Section

Standard 16. Safety

- | | Yes | No | N/A |
|--|-------------------------------------|-------------------------------------|--------------------------|
| 16.1 Does the laboratory have and follow a documented environmental health and safety program that includes, at a minimum, the following: | <input type="checkbox"/> | <input checked="" type="checkbox"/> | <input type="checkbox"/> |
| 16.1.1 A bloodborne pathogen and chemical hygiene plan? | <input checked="" type="checkbox"/> | <input type="checkbox"/> | <input type="checkbox"/> |
| 16.1.2 Documented training on the bloodborne pathogen and chemical hygiene plan? | <input checked="" type="checkbox"/> | <input type="checkbox"/> | <input type="checkbox"/> |
| 16.2 Has the laboratory's environmental health and safety program been reviewed annually? | <input checked="" type="checkbox"/> | <input type="checkbox"/> | <input type="checkbox"/> |
| a. Has such review been documented? | <input checked="" type="checkbox"/> | <input type="checkbox"/> | <input type="checkbox"/> |

[Click Here For Discussion](#)

Comment

Standard 16.1 - See Finding Section

STANDARD 17. Outsourcing

| | | Yes | No | N/A |
|---------------|---|------------------------------|-----------------------------|-------------------------------------|
| 17.1 | Has the vendor laboratory complied with the FBI Quality Assurance Standards for Forensic DNA Testing Laboratories and the accreditation requirements of federal law? | <input type="checkbox"/> | <input type="checkbox"/> | <input checked="" type="checkbox"/> |
| 17.1.1 | Has the NDIS laboratory that outsources DNA sample(s) for entry into CODIS required and maintained the following documentation from the vendor laboratory: | <input type="checkbox"/> | <input type="checkbox"/> | <input checked="" type="checkbox"/> |
| | a. Compliance with the FBI Quality Assurance Standards for Forensic DNA Testing Laboratories? | Yes <input type="checkbox"/> | No <input type="checkbox"/> | |
| | b. Compliance with the accreditation requirements of federal law? | Yes <input type="checkbox"/> | No <input type="checkbox"/> | |
| 17.2 | Except as provided in Standard 17.2.1, since the laboratory's last external audit, did the NDIS laboratory's technical leader document and maintain the approval of the technical specifications of the outsourcing agreement before it was awarded? | <input type="checkbox"/> | <input type="checkbox"/> | <input checked="" type="checkbox"/> |
| 17.2.1 | For a vendor laboratory that is performing forensic DNA analysis for a law enforcement agency or entity other than the NDIS laboratory, was documented approval obtained by the vendor laboratory from the technical leader of the NDIS laboratory, accepting ownership of the DNA data generated, prior to the initiation of analysis? | <input type="checkbox"/> | <input type="checkbox"/> | <input checked="" type="checkbox"/> |
| 17.3 | Did the NDIS laboratory accept, upload to, or search in CODIS, profiles generated by a vendor laboratory? | <input type="checkbox"/> | <input type="checkbox"/> | <input checked="" type="checkbox"/> |

- a. Prior to the NDIS laboratory's uploading or accepting data to upload or search in CODIS from any vendor laboratory or agency, did the technical leader of the NDIS laboratory document the prior approval of the technical specifications of the outsourcing agreement and/or document the approval of acceptance of ownership of the DNA data?
- 17.4** Does the NDIS laboratory have and follow a procedure to verify the integrity of the data received from a vendor laboratory through the performance of a technical review?
- 17.5** Prior to the upload or search of the data generated by the vendor laboratory to SDIS, did the NDIS laboratory perform a technical review of the vendor laboratory's data?
 - a. Was the technical review performed by an NDIS laboratory-employed analyst or technical reviewer who is qualified, or was previously qualified, in the technology, platform, and typing amplification test kit used to generate the data and who participates in the NDIS laboratory's proficiency-test program?
- 17.5.1** Do the technical review procedures include, at a minimum, the following elements:
 - 17.5.1.1** A review of all DNA types to verify that they are supported by the raw and/or analyzed data? (electropherograms or images)
 - 17.5.1.2** A review of all associated controls, internal lane standards and allelic ladders to verify that the expected results were obtained?
 - 17.5.1.3** A review of the final report (if provided) to verify:
 - a. That the results/conclusions are supported by the data?

Yes No
 - b. That each tested item (or its probative fraction) submitted to the vendor laboratory is addressed?

Yes No
 - 17.5.1.4** Verification of the DNA types, eligibility, and the correct specimen category for entry into CODIS?
- 17.6** For an on site visit:
 - a. Does the NDIS laboratory have and follow a procedure for performing an on-site visit?

- b. Does the procedure include, at a minimum, the following elements?
- 17.6.1** A documented on-site visit prior to the initiation of analysis?
 - 17.6.1.1** Has the on-site visit been performed by either the technical leader or a designated employee of the NDIS laboratory who is a qualified or previously qualified analyst in the technology, platform, and typing amplification test kit used to generate the DNA data?
- 17.6.2** If the NDIS laboratory's outsourcing agreement extended beyond one year, was an annual on-site visit conducted?
 - 17.6.2.1** If an on-site visit conducted by another NDIS laboratory was used by the NDIS laboratory, did the technical leader document the review and acceptance of that on-site visit?

[Click Here For Discussion](#)

Standard 17 and all sub-standards were marked N/A as the laboratory does not outsource casework samples.

Appendix A: Findings and Responses

Findings:

Standard 3.1.1 Is the quality system documented in a manual that includes or references the following elements:

Standard 3.1.1.4 Facilities?

Objective Proof for the Finding

The Quality Manual, version 3 (Revised 04/2007) that was in effect until April 1, 2011, did not define, establish, or reference the laboratory's practices or procedures for laboratory security and its approach for maintaining the integrity of DNA analyses and evidence examination.

The Technical Leader provided the audit team with a memo dated 7/13/2011 stating that he had conducted annual reviews in 2009 and 2010 of the quality system as required by standard 3.3 and that the review encompassed a review and approval of all written analytical DNA procedures as well as elements required in standards 5.2.5.1 and 5.2.5.2, and no protocol modifications were made. This requirement was from the July 2009 QAS and was not updated in a laboratory manual until April 2011.

Standard 3.1.1 Is the quality system documented in a manual that includes or references the following elements:

Standard 3.1.1.15 Outsourcing?

Objective Proof for the Finding:

The Quality Manual, version 3 (Revised 04/2007) that was in effect until April 1, 2011, does not define, establish, or reference the laboratory's procedures for outsourcing samples and ensuring the integrity of those samples.

The Technical Leader provided the audit team with a memo dated 7/13/2011 stating that he had conducted annual reviews in 2009 and 2010 of the quality system as required by standard 3.3 and that the review encompassed a review and approval of all written analytical DNA procedures as well as elements required in standards 5.2.5.1 and 5.2.5.2, and no protocol modifications were made. This requirement was from the July 2009 QAS and was not updated in a laboratory manual until April 2011.

Appendix A: Findings and Responses

Findings:

Standard 4.1 Does the laboratory have:

Standard 4.1.2 A Technical leader who is accountable for the technical operations?

Objective Proof for the Finding:

The discussion under this standard states, 'Standard 5.2.3.1 and its subcategories must be satisfied in order to demonstrate that the Technical Leader is accountable for the technical operations.'

The Technical Leader's job description defined in Appendix VIII - DNA Unit Job Functions that was in effect until April 1, 2011, does not state that the Technical Leader has the authority to initiate, suspend, and resume DNA database operations for the laboratory or an individual. Standard 5.2.3.1.2 (a subcategory of 5.2.3.1) is marked NO and therefore this standard is also marked NO.

This requirement was from the July 2009 QAS and was not updated in a laboratory manual until April 2011. See findings under standards 3.1.1 and 5.2.3.1).

Standard 4.1 Does the laboratory have:

Standard 4.1.3 A casework CODIS administrator who is accountable for CODIS on-site at each individual laboratory facility using CODIS?

Objective Proof for the Finding:

The discussion under this standards states, 'Standard 5.3.5 must be satisfied in order to demonstrate that the casework CODIS administrator is accountable for CODIS operations on-site at each individual laboratory facility using CODIS.' The CODIS administrator's job description defined in Appendix VIII - DNA Unit Job Functions that was in effect until April 1, 2011, does not state that the CODIS Administrator is authorized to terminate an analyst's or the laboratory's participation in CODIS until the reliability and security of the computer data can be assured if an issue with the data is identified.

Standard 5.3.5 is marked NO and therefore this standard is also marked NO.

This requirement was from the July 2009 QAS and was not updated in a laboratory manual until April 2011. See findings under standards 3.1.1 and 5.3.5).

Appendix A: Findings and Responses

Findings:

Standard 4.1 Does the laboratory have:

Standard 4.1.5 Documentation that specifies the responsibility, authority, and interrelation of all personnel who manage, perform, or verify work affecting the validity of the DNA analysis?

Objective Proof for the Finding:

The laboratory has an organizational chart (June 2011) however the chart does not reference specific personnel by name with their specific assignments (Technical Leader, casework or CODIS administrator) or reference the specific position assignments (Technical Leader or casework CODIS administrator).

The DNA Job Unit Functions Appendix VIII revised 12/01/2008 includes an organizational chart that lists position titles; however it is outdated and it doesn't reflect that there is only one DNA Technical Leader (as of August 2009) for the laboratory and still shows a separate DNA Technical Leader for the mtDNA section. The DNA Section Job Functions, Appendix 7, issued April 2011 does reflect only one DNA Technical Leader.

The Technical Leader provided the audit team with a memo dated 7/13/2011 stating that he had conducted annual reviews in 2009 and 2010 of the quality system as required by standard 3.3 and that the review encompassed a review and approval of all written analytical DNA procedures as well as elements required in standards 5.2.5.1 and 5.2.5.2, and no protocol modifications were made. The DNA Job Unit Functions was not updated in a laboratory manual until April 2011.

Standard 5.1 Do laboratory personnel have the education, training, and experience commensurate with the examination and testimony provided?

Note:

To successfully satisfy Standard 5.1, compliance must be demonstrated with all of the subcategories of Standard 5. There is no separate finding against this standard, only that the laboratory was not in compliance for Standards 5.1.1, 5.1.2.1.a, 5.1.2.2, 5.1.3.2.a, 5.2, 5.2.1, 5.2.1.b, 5.2.1.b.4, 5.2.1.1, 5.2.3.1.2, 5.2.3.2.1, 5.2.3.2.2, 5.2.3.2.3, 5.2.3.2.4, 5.2.3.2.5, 5.2.5.1, 5.3.4, 5.3.4.2, 5.3.4.3, 5.3.4.4, 5.3.4.5, 5.3.5, and 5.4.1.3.b.

Appendix A: Findings and Responses

Findings:

Standard 5.1.1 Does the laboratory have written job descriptions for all personnel to include responsibilities, duties, and skills?

Objective Proof for the Finding:

The DNA Technical Leader and casework CODIS Administrator job descriptions do not specify all of the responsibilities, duties and skills for this position as described in these standards.

Note:

The following was a finding from the last external audit conducted December 2008: While the laboratory did provide the assessment team with general job descriptions for laboratory scientists and technicians and did specify and document the responsibility, authority and interrelation of the Technical Leader and CODIS administrator, they did not provide the assessors with job descriptions, responsibilities and relevant authority of nuclear DNA analysts and nuclear DNA technicians.

This finding was addressed by the laboratory and submitted to NDIS but was not included in the laboratory manuals until April 2011, in the DNA Section Job Functions Appendix 7 issued April 2011. The Technical Leader provided the audit team with a memo stating that he had conducted annual reviews in 2009 and 2010 of all processes and that no changes were needed. Please refer to Standard 3.1.1.

Standard 5.1.2.1 Does the training program contain at a minimum the following components:

Standard 5.1.2.1.a A training manual that covers all applicable DNA analytical procedures that the analyst/technician will perform?

Objective Proof for the Finding:

A laboratory's training must include all methodologies that an analyst will perform in casework analysis. The laboratory's Training Manual (2010, rev. 004) does not teach and assess mtDNA extraction, amplification and analysis procedures.

Appendix A: Findings and Responses

Findings:

Standard 5.1.2.2 Does the laboratory's training program teach and assess the technical skills and knowledge required to perform DNA analysis and include, at a minimum, the following?

Objective Proof for the Finding:

As per the discussion under this standard, a laboratory's training program must teach and assess the skills and knowledge required to achieve the minimum standards of competence and good laboratory practice in a specific area of work. The laboratory's training program does include all relevant aspects of a comprehensive training program for autosomal and Y STR analysis, however interviews with the staff revealed that it is not effective in all areas. In addition, there was a lack of documentation in paternity training in the analysts' training binders

One analyst had performed DNA analysis on a criminal paternity case and when interviewed regarding that statistical application of paternity statistics he said he was not comfortable with conducting the statistical analysis or testifying to these statistics; he did not recall being trained in paternity statistics. When this was brought to the Technical Leader's attention he stated that the analyst would not be allowed to testify to the paternity statistics and that the laboratory would have the other analyst that signed the report testify instead. The Technical Leader said that both analysts assume responsibility for the content of the report and therefore both can testify to the contents. (see findings under standard 11.2.9).

Additionally, several analysts were asked about theta corrections and/or confidence intervals. At least three individuals interviewed did not have an understanding as to what theta and/or a confidence interval was or how it is applied.

Standard 5.1.3.2 For the review of scientific literature:

Standard 5.1.3.2.a Does the laboratory have a program, approved by the technical leader, for the annual review of scientific literature that documents the ongoing reading of scientific literature?

Objective Proof for the Finding:

The laboratory's DNA SOP-1, section 1.14.3 states, "All DNA analysts will attend at least one continuing education unit (8 hours) and read at least one relevant scientific paper each year as outlined in the FBI QAS (Standard 5). This record will be documented in the analyst's personnel file." Most analysts are only documenting the reading of one scientific article per year, which is not sufficient to stay abreast of new developments and issues in the field of DNA analysis. The Technical Leader stated that most analysts are reading more than one article per year, but the audit team was given no documentation to support this. Additionally, there is no documentation to demonstrate that the Technical Leader is approving the reading of scientific literature.

Appendix A: Findings and Responses

Findings:

Standard 5.2 Does the technical leader satisfy the requirements for degree/education, experience, and duties listed in Standards 5.2.1 through 5.2.4.1?

Standard 5.2.1 Does the technical leader of the laboratory meet or exceed the following degree/educational requirements?

Standard 5.2.1.b Twelve semester hours or equivalent credit hours including a combination of graduate and undergraduate course work or classes covering the following subject areas:

Standard 5.2.1.b.4 Statistics or population genetics?

Standard 5.2.1.1 Of the 12 semester or equivalent credit hours required, do they include at least one graduate-level course registering 3 or more semester or equivalent credit hours?

The following were findings from the 12/2008 external audit of this laboratory with reference to the mtDNA Technical Leader, Mary Beth Raffin. This Technical Leader was not removed from her responsibilities until August 2009. The audit team was given no documentation to demonstrate that these findings were adequately addressed. Therefore these were included as findings in this document.

Objective Proof for the Finding:

The individual assigned as Technical Leader to the mtDNA analysis section has not completed a course in statistics and/or population genetics. Additionally, the individual assigned as Technical Leader has no graduate course work in biochemistry, genetics, molecular biology or statistics/population genetics.

Through interviews and case file discussions the Technical manager/leader could not demonstrate general knowledge and oversight of the mtDNA program sufficient to ensure the laboratory is following standards and written protocols. An analyst was required to provide a review of laboratory case file documentation as some points could not be explained by the technical leader.

Interviews with laboratory staff and the assigned Technical Leader indicate that the Technical Leader has not kept up to date with mtDNA analysis through relevant review of mtDNA literature, internal laboratory validation review or continuing education in the area of mtDNA analysis to a point sufficient to: manage technical operations of the laboratory, evaluate methods used by the laboratory, propose new or modified analytical procedures to be used by the examiners or provide technical support in problem solving of analytical procedures.

The Technical Leader also has not been given oversight of corrective action. Corrective actions taken in response to an incorrect primer sequence (critical reagent) were directed by laboratory analysts and reported to the laboratory Quality Manager. The Technical Leader was only notified once the action plan was completed.

Appendix A: Findings and Responses

Findings:

Standard 5.2.3 Does the technical leader of the laboratory have responsibility for the following:

5.2.3.1 Does the technical leader have the following general duties and authority:

Standard 5.2.3.1.2 Authority to initiate, suspend, and resume DNA analytical operations for the laboratory or an individual?

Objective Proof for the Finding:

The Technical Leader's job description listed in the DNA Job Unit Functions Appendix VIII revised 12/01/2008 does not state that the Technical Leader has the authority to initiate, suspend, and resume DNA operations for the laboratory or an individual. This was not updated in the laboratory manuals until April 2011, in the DNA Section Job Functions Appendix 7 issued April 2011. This requirement was new to the July 2009 QAS and was not updated in a laboratory manual until April 2011.

Standard 5.2.3 Does the technical leader of the laboratory have responsibility for the following:

Standard 5.2.3.2 Does the technical leader perform the following specific responsibilities:

Standard 5.2.3.2.1 Evaluate and document approval of all validations and methods used by the laboratory and propose new or modified analytical procedures to be used by analysts?

Objective Proof for the Finding:

The Technical Leader's job description listed in the DNA Job Unit Functions Appendix VIII revised 12/01/2008 does not state that the Technical Leader is responsible for evaluating and documenting approval of all validations and methods used by the laboratory. This was not updated in the laboratory manuals until April 2011, in the DNA Section Job Functions Appendix 7 issued April 2011. This requirement was modified in the July 2009 QAS and was not updated in a laboratory manual until April 2011.

Additionally, the Technical Leader assumed responsibility as the mtDNA Technical Leader in August 2009 and there is no objective proof that he approved all validations and methods used by the mtDNA section.

Appendix A: Findings and Responses

Findings:

Standard 5.2.3 Does the technical leader of the laboratory have responsibility for the following:

Standard 5.2.3.2 Does the technical leader perform the following specific responsibilities:

Standard 5.2.3.2.2 Review and document the review of the academic transcripts and training records for newly qualified analysts and approve their qualifications prior to their conducting independent casework analysis?

Objective Proof for the Finding:

The Technical Leader's job description listed in the DNA Job Unit Functions Appendix VIII revised 12/01/2008 does not state that the Technical Leader is responsible for reviewing and documenting the review of the academic transcripts and training records for newly qualified analysts and approve their qualifications prior to their conducting independent casework analysis. This was not addressed in the updated laboratory manual issued April 2011, in the DNA Section Job Functions Appendix 7 issued April 2011. This requirement was new to the July 2009 QAS and was not updated in a laboratory manual until April 2011.

Additionally, several analysts were qualified and began casework on or after 7/1/2009 and the Technical Leader did not approve their academic transcripts until 6/27/2011.

Standard 5.2.3 Does the technical leader of the laboratory have responsibility for the following:

Standard 5.2.3.2 Does the technical leader perform the following specific responsibilities:

Standard 5.2.3.2.3 Approve the technical specifications for outsourcing agreements?

Objective Proof for the Finding:

The Technical Leader's job description listed in the DNA Job Unit Functions Appendix VIII revised 12/01/2008 does not state that the Technical Leader is responsible for approving the technical specifications for outsourcing agreements. This was not addressed in the updated laboratory manual issued April 2011, in the DNA Section Job Functions Appendix 7 issued April 2011. This requirement was new to the July 2009 QAS and was not updated in a laboratory manual until April 2011.

Appendix A: Findings and Responses

Findings:

Standard 5.2.3 Does the technical leader of the laboratory have responsibility for the following:

Standard 5.2.3.2 Does the technical leader perform the following specific responsibilities:

Standard 5.2.3.2.4 Review and document the review of internal and external DNA audit documents and, if applicable, approve corrective action(s).

Objective Proof for the Finding:

The Technical Leader's job description listed in the DNA Job Unit Functions Appendix VIII revised 12/01/2008 does not state that the Technical Leader is responsible for reviewing and documenting the review of internal and external DNA audit documents and, if applicable, approve corrective action(s). This was not updated in the laboratory manuals until April 2011, in the DNA Section Job Functions Appendix 7 issued April 2011. The 2011 revisions state the Technical Leader's responsibility is for oversight and approval of validation, internal and external audits. The new manual does not state that the Technical Leader must approve corrective action(s). This requirement was new to the July 2009 QAS and was not updated in a laboratory manual until April 2011.

Standard 5.2.3 Does the technical leader of the laboratory have responsibility for the following:

Standard 5.2.3.2 Does the technical leader perform the following specific responsibilities:

Standard 5.2.3.2.5 Review annually the procedures of the laboratory and document such review?

Objective Proof for the Finding:

The Technical Leader's job description listed in the DNA Job Unit Functions Appendix VIII revised 12/01/2008 does not state that the Technical Leader is responsible for annually reviewing the procedures of the laboratory and documenting these reviews. This was not addressed in the updated laboratory manual issued April 2011, in the DNA Section Job Functions Appendix 7 issue April 2011. This requirement was new to the July 2009 QAS and was not updated in a laboratory manual until April 2011.

Appendix A: Findings and Responses

Findings:

Standard 5.2.5 Did each technical leader appointed or hired on or after July 1, 2009, document a review of the following:

Standard 5.2.5.1 Validation studies and methodologies currently used by the laboratory?

Objective Proof for the Finding:

The Technical Leader was appointed as the mtDNA Technical Leader in August 2009. There is no documentation to demonstrate that he reviewed the validation studies and methodologies currently used by the mtDNA section. This requirement was new to the July 2009 QAS and as of the date of this audit had not been conducted.

Standard 5.3.4 Is the casework CODIS administrator responsible for the following:

Standard 5.3.4.2 Scheduling and documenting the CODIS computer training of casework analysts?

Objective Proof for the Finding:

The casework CODIS administrator's job description listed in the DNA Job Unit Functions Appendix VIII revised 12/01/2008 does not state that the casework CODIS administrator is responsible for scheduling and documenting the CODIS computer training of casework analysts. This was not updated in the laboratory manuals until April 2011, in the DNA Section Job Functions Appendix 7 issued April 2011. This requirement for casework CODIS Administrators was new to the July 2009 QAS and was not updated in a laboratory manual until April 2011.

Standard 5.3.4 Is the casework CODIS administrator responsible for the following:

Standard 5.3.4.3 Assuring that the security of data stored in CODIS is in accordance with state and/or federal law and NDIS operational procedures?

Objective Proof for the Finding:

The casework CODIS administrator's job description listed in the DNA Job Unit Functions Appendix VIII revised 12/01/2008 does not state that the casework CODIS administrator is responsible for assuring that the security of data stored in CODIS is in accordance with state and/or federal law and NDIS operational procedures. This was NOT updated in the laboratory manuals until April 2011, in the DNA Section Job Functions Appendix 7 issued April 2011. This requirement for casework CODIS Administrators was new to the July 2009 QAS and was not updated in a laboratory manual until April 2011.

Appendix A: Findings and Responses

Findings:

Standard 5.3.4 Is the casework CODIS administrator responsible for the following:

Standard 5.3.4.4 Assuring that the quality of data stored in CODIS is in accordance with state and/or federal law and NDIS operational procedures?

Objective Proof for the Finding:

The casework CODIS administrator's job description listed in the DNA Job Unit Functions Appendix VIII revised 12/01/2008 does not state that the casework CODIS administrator is responsible for assuring that the quality of data stored in CODIS is in accordance with state and/or federal law and NDIS operational procedures. This was NOT updated in the laboratory manuals until April 2011, in the DNA Section Job Functions Appendix 7 issued April 2011. This requirement for casework CODIS Administrators was new to the July 2009 QAS and was not updated in a laboratory manual until April 2011.

Standard 5.3.4 Is the casework CODIS administrator responsible for the following:

Standard 5.3.4.5 Assuring that matches are dispositioned in accordance with NDIS operational procedures?

Objective Proof for the Finding:

The casework CODIS administrator's job description listed in the DNA Job Unit Functions Appendix VIII revised 12/01/2008 does not state that the casework CODIS administrator is responsible for assuring that matches are dispositioned in accordance with NDIS operational procedures. This was not updated in the laboratory manuals until April 2011, in the DNA Section Job Functions Appendix 7 issued April 2011. This requirement for casework CODIS Administrators was new to the July 2009 QAS and was not updated in a laboratory manual until April 2011.

Standard 5.3.5 Is the CODIS administrator authorized to terminate an analyst's or the laboratory's participation in CODIS until the reliability and security of the computer data can be assured if an issue with the data is identified?

Objective Proof for the Finding:

The casework CODIS administrator's job description listed in the DNA Job Unit Functions Appendix VIII revised 12/01/2008 does not state that the casework CODIS administrator is authorized to terminate an analyst's or the laboratory's participation in CODIS until the reliability and security of the computer data can be assured if an issue with the data is identified. This was not updated in the laboratory manuals until April 2011, in the DNA Section Job Functions Appendix 7 issued April 2011. Please refer to the discussion under 5.1.1. This requirement for casework CODIS Administrators was new to the July 2009 QAS and was not updated in a laboratory manual until April 2011.

Appendix A: Findings and Responses

Findings:

Standard 5.4.1.3 For individuals who have completed coursework with titles other than those listed in Standard 5.4.1:

Standard 5.4.1.3.b. Has the technical leader documented his or her approval of compliance with this Standard?

Objective Proof for the Finding:

During the audit, one analyst could not demonstrate that a genetics course (a minimum of 3 semester hours) was taken, since the coursework was not entitled 'genetics'. The Technical Leader signed off approval of this individual's transcripts on 6/27/2011 and records show he went on line with casework 6/25/10; however there was no documentation such as a transcript, syllabus, letter from the instructor, or other documentation to support the approval of this course content for genetics. Two days after the audit closeout, the audit team was provided with a letter from the dean of the college stating that the genetics lectures and homework from courses taken for his degree comprised an integral component of those classes and is greater than three credit-hours of genetics class instruction at the graduate level. This was used as documentation to support the genetics course for the analyst; however the Technical Leader did not make any attempts to obtain this objective proof prior to this audit or his approval of the analyst's transcripts.

Standard 7.1 Does the laboratory have and follow a documented evidence control system to ensure the integrity of physical evidence?

Note:

To successfully satisfy Standard 7.1, compliance must be demonstrated with all of the subcategories of Standard 7. There is no separate finding against this standard, only that the laboratory was not in compliance for Standards 7.1.1.b and 7.3.

Appendix A: Findings and Responses

Findings:

Standard 7.1.1 For evidence and sample identification:

Standard 7.1.1.b Does the laboratory clearly define what constitutes evidence and what constitutes work product?

Objective Proof for the Finding:

The Technical Leader issued a memorandum on 12/02/06 which states "...genomic DNA extracts are considered to be work product not evidence". There is no documentation that clearly defines what constitutes evidence and what constitutes work product. Additionally this 'policy' was not incorporated into the 2008 or 2011 procedures manuals.

The annual review, which the Technical Leader stated he conducted, did not result in changes to the standard operating procedures to reflect this practice, and in fact, the memo provided by the Technical Leader (dated 7/13/2011) specifically stated, 'This Quality system review for 2009 and 2010 encompassed a review (and approval) by the Technical Leader of all written analytical DNA procedures as well as the elements required in Standard 5.2.5.1 & 5.2.5.2. As a result, no protocol modifications were made.'

Standard 7.3 Does the laboratory have and follow documented policies for the disposition of evidence and sample consumption?

Objective Proof for the Finding:

The laboratory does not have a documented policy on sample consumption.

Note: The Technical Leader issued a Consumption of Evidence policy (Document ID DNA WI-23) on 7/11/11 which adequately addressed the finding; however there was nothing in place prior to July 11, 2011.

Standard 8.1 Does the laboratory use validated methods for DNA analyses?

Note:

To successfully satisfy Standard 8.1, compliance must be demonstrated with all of the subcategories of Standard 8. There is no separate finding against this standard, only that the laboratory was not in compliance for Standards 8.3.1.b, 8.3.1.b.1, 8.3.1.b.5, 8.5 and 8.6.

Appendix A: Findings and Responses

Findings:

Standard 8.3.1 For Internal Validation Studies:

Standard 8.3.1.b Have all internal validation studies conducted on or after July 1, 2009, included, as applicable:

Standard 8.3.1.b.1 Known and non probative evidence samples or mock evidence samples?

Standard 8.3.1.b.5 Contamination assessment?

Objective Proof for the Finding:

The laboratory conducted internal validation studies for the Quantifiler® Duo Kit which were approved by the Technical Leader on 8/27/2009. This validation study encompassed precision/stability, reproducibility/sensitivity, mixtures, degradation and specificity. The studies were conducted using known samples and various sources of animal DNA. No non-probative evidence or mock evidence samples were utilized in the validation study. Additionally, a contamination assessment was not included in the validation summary or noted anywhere in the validation data provided for review.

Standard 8.5 Have modified procedures been evaluated by comparison with the original procedures using similar DNA samples prior to their incorporation into casework applications?

Objective Proof for the Finding:

A material modification was conducted to allow the diluted standards (buffer and concentrated standard) from the Quantifiler® Duo Kit to be used with the primers and reaction mix from the Quantifiler® Human Kit. Analysts also have the option of using the Quantifiler® Human Kit to quantitate known/reference samples. This modification allowed the laboratory to generate one set of diluted standards to be used for both kits rather than generating a diluted set of standards for the Quantifiler® Duo Kit and one for the Quantifiler® Human Kit. A comparison study was conducted and summarized in a memo dated 12/21/2009. The audit team requested supporting data and analysis documentation to support the summary conclusion that the use of reagents from the Quantifiler® Duo kit gave comparable results when compared to those in Quantifiler® Human kit.

No documentation regarding the approval date by the Technical Leader could be located or provided to the audit team except the material modification process that was added to the SOP 3.3.5 dated 4/1/11. Therefore, although a comparison study was conducted, no evidence of the evaluation of the data and approval by the Technical Leader was provided prior to incorporation into casework (Also see finding under Standard 9.1).

Appendix A: Findings and Responses

Findings:

Continued from previous page:

Standard 8.5 Have modified procedures been evaluated by comparison with the original procedures using similar DNA samples prior to their incorporation into casework applications?

Objective Proof for the Finding - continued:

Data was provided to the audit team by the analyst that performed the validation. A comparison study was performed and the summary stated the modification was comparable to current methods; however, no supportive data was found in the binder and additional documentation provided to the audit team did not contain the specific data used in the evaluation. The analyst stated that the Y-intercept was used in the analysis, but the data provided to the audit team did not include Y-intercept information. Additionally, the data set had no identifiers (date/title) that could be used to link the data set to this particular material modification. Please see finding under standard 8.3.1.a.

Additionally, a modification of the expiration date for a diluted set of quantitation standards from the Quantifiler® Duo kit was evaluated. The manufacturer states that the diluted set of standards can be stored up to two weeks at 2-8°C and longer storage is not recommended.

A comparison study was conducted at the laboratory and it was determined that a two month storage period did not have a negative effect on the quality of the standard curve. Data was generated, but no statistical analysis was conducted to support the conclusion that the data sets were comparable. In SOP 3.0 Quantitation, the SOP states that "The diluted DNA quantitation standards may be stored for longer than 2 weeks at 2 to 8°C..." No set expiration date was documented in the SOP and appears to be infinite and at the analyst's discretion and is not supported by the material modification conducted by the laboratory.

Appendix A: Findings and Responses

Findings:

Standard 8.6 Has the laboratory evaluated each additional or modified critical instrument by conducting a performance check prior to its use in casework?

Objective Proof for the Finding:

If a laboratory currently uses one instrument and adds other instruments of the same model number, it is necessary to conduct a performance check on the new instruments. This laboratory currently utilizes injection times on the Applied Biosystems' 3130xl Genetic Analyzer ranging from 2-60 seconds. Using the Applied Biosystems' Identifiler® kit, performance checks were conducted on the new AB 3130xl -3 and 3130xl -4 instruments. The only injection time used for the performance check on the 3130xl -3 instrument was 10 seconds and the only injection times used for the performance check on the 3130xl -4 instrument were 5 and 10 seconds.

It was observed during casework reviews that injection times generally ranged from 5-40 seconds. The injection time has a direct impact on the sensitivity and ability to detect an analyte, by not assessing the full dynamic range of the injection times used by the laboratory, the staff would be unable to determine if these new instruments were working as expected in these injection ranges. The performance checks do not adequately demonstrate that the new instruments perform as expected in the dynamic range of injection times defined by the laboratory. Additionally, increased injection time is not specifically noted as an enhancement techniques.

Additionally, the laboratory conducts Y STR testing using the PowerPlex® Y System on the AB 3130xl genetic analyzers. The performance check did not include running the PowerPlex® Y System. Given that the Identifiler® and PowerPlex® Y kits have different amplification and dye chemistries, operational parameters and sensitivity, the performance check would not adequately assess that the PowerPlex® Y kit was performing as expected. (see finding under Standard 10.2.1.7).

Appendix A: Findings and Responses

Findings:

Standard 9.1 Does the laboratory have and follow written analytical procedures approved by the technical leader?

Objective Proof for the Finding:

A material modification was conducted to allow the diluted standards (buffer and concentrated standard) from the Quantifiler® Duo Kit to be used with the primers and reaction mix from the Quantifiler® Human Kit. Analysts also have the option of using the Quantifiler® Human Kit to quantitate known/reference samples. This modification allowed the laboratory to generate one set of diluted standards to be used for both kits rather than generating a diluted set of standards for the Quantifiler® Duo Kit and one for the Quantifiler® Human Kit. A comparison study was conducted and summarized in a memo dated 12/21/2009. The audit team requested supporting data and analysis documentation to support the summary conclusion that the use of reagents from the Quantifiler® Duo kit gave comparable results when compared to those in Quantifiler® Human kit.

Documentation of the approval of the modification by the Technical Leader consisted of initials at the top of the Quantifiler Human & Y DNA Worksheets but did not include a date indicating when the method was officially approved for use. No documentation regarding the approval date by the Technical Leader could be located or provided to the audit team except the material modification process that was added to the SOP 3.3.5 dated 4/1/11. Prior to 4/1/11, this modification did not appear in any DNA SOP. (Also see finding under Standard 8.3.1.a).

Standard 9.1.a Are the laboratory's standard operating procedures reviewed annually by the technical leader, and is this review documented?

Objective Proof for the Finding:

The laboratory could not provide documentation of the Technical Leader's annual review of the analytical procedures. A memo was drafted by the Technical Leader during the audit stating that the Technical Leader conducted the quality review of the DNA process pursuant to "Standard 3.3"; however it did not mention or include analytical procedures. The team was provided with a modified memo dated 7/13/2011 stating that this review included all written analytical DNA procedures as well as the elements required in standards 5.2.5.1 and 5.2.5.2.

The analytical procedures provided to the audit team had a revision date of 12-01-2008.

Appendix A: Findings and Responses

Findings:

Continued from previous page:

Standard 9.1.a Are the laboratory's standard operating procedures reviewed annually by the technical leader, and is this review documented?

Objective Proof for the Finding - continued:

Although "Work Instructions" were provided for the procedure, the standard operating procedures were not up-dated with currently used procedures until April 1, 2011. The Technical Leader approved the Quantifiler® Duo DNA Quantification Kit "Work Instructions" 8/27/2009 to allow the diluted standards (buffer and concentrated standard) from the Quantifiler® Duo Kit to be used with the primers and reaction mix from the Quantifiler® Human Kit. The audit team requested supporting data and analysis documentation to support the summary conclusion that the use of reagents from the Quantifiler® Duo kit gave comparable results when compared to those in Quantifiler® Human kit to allow the Quantifiler® Duo DNA Quantification Kit dilution buffer to dilute the Quantifiler® Human DNA Quantification Kit standard DNA. The "Work Instruction" provided to the audit team does not include this material modification.

The annual review, which the Technical Leader stated he conducted, did not result in changes to the standard operating procedures to reflect the current practices employed by the laboratory and in fact, the memo provided by the Technical Leader (referenced above - dated 7/13/2011) specifically stated, 'This Quality system review for 2009 and 2010 encompassed a review (and approval) by the Technical Leader of all written analytical DNA procedures as well as the elements required in Standard 5.2.5.1 & 5.2.5.2. As a result, no protocol modifications were made.' Protocol modifications were made in practice, but they were not identified in the annual reviews or included as revisions to the written protocols.

The Technical Leader assumed responsibility for the mtDNA operation in August 2009; however there is no documentation to demonstrate that he has ever reviewed the procedures.

During interviews, the CODIS manager stated that the laboratory has the NDIS procedures for CODIS + Mito; however, these do not address state level defined search parameters for missing persons, specific instructions regarding building pedigree trees, how to enter metadata, and laboratory defined match criteria for missing persons at the state level. He stated that in practice, he was not familiar with some of the parameters and thought the laboratory was "probably" using the default settings. One analyst interviewed by the audit team stated, the mito unit's practice has been to NOT enter metadata and there has not been a mito match since the inception of CODIS + Mito at their laboratory. The audit team requested NDIS level searches; however these were never provided.

Appendix A: Findings and Responses

Findings:

Standard 9.1.1 Does the laboratory have a documented standard operating procedure for each analytical method used?

Objective Proof for the Finding:

The laboratory does not have a CODIS Users Manual for CODIS+Mito (ver. 6.1).

The laboratory's Quality Manual 5.0 Technical Procedures Section (Revision 3, 04/2007) states:

5.4 When a need for changes or additions to the technical procedures is identified, that modification will occur only after the technical issue has been thoroughly investigated, evaluated, validated and documented.

B4 - A notation will be made in the front of the technical manual listing the modifications, date of modification, and reason (see attached form QR 5-1).

B5 - All replaced or modified material will be collected by the Quality Manager and maintained in the quality records. The revision # and date shall appear on each page the new or modified procedure.

The laboratory's Quality Manual 5.0 Technical Procedures Section (Revision 3, 04/2007) does not include any information concerning the existence of "Work Instructions", how "Work Instructions" relate to the laboratory's Standard Operating Protocols or methods for up-date and revision of "Work Instructions". The Standard Operating Protocols provided to the audit team had revision dates of 12/1/2008. The laboratory provided "Work Instructions" for the Biomek® 2000 robotic workstation using the Promega DNA IQ™ extraction chemistry (approved 7/28/2009) and the Quantifiler® Duo DNA Quantification Kit (approved 8/27/2009) as the analytical procedures for these methods. These Work Instructions provided to the audit team for these methods contained the Technical Leader's initials and the date 8/27/09. The "Work Instructions" did not contain revision dates or documentation of annual review. The laboratory did not provide documentation of a notation made in the front of the technical manual listing the modifications, date of modification and reason or a copy of the QR 5-1 form.

The Technical Leader approved the Quantifiler® Duo DNA Quantification Kit "Work Instructions" 8/27/2009. A material modification was validated 12/2009 to allow the Quantifiler® Duo DNA Quantification Kit dilution buffer to dilute the Quantifiler® Human DNA Quantification Kit standard DNA. The "Work Instruction" provided to the audit team does not include this material modification.

As noted above, the annual review did not result in changes to the standard operating procedures to reflect the current practices employed by the laboratory (to include work instructions) in fact, the memo provided by the technical leader (referenced above - dated 7/13/2011) specifically stated, 'This Quality system review for 2009 and 2010 encompassed a review (and approval) by the Technical Leader of all written analytical DNA procedures as well as the elements required in Standard 5.2.5.1 & 5.2.5.2. As a result, no protocol modifications were made.' Protocol modifications were made in practice, but they were not identified in the annual reviews or included as revisions to the written protocols until April 2011.

Appendix A: Findings and Responses

Findings:

Standard 9.1.1.a Do the analytical procedures specify reagents, sample preparation, extraction methods, equipment, and controls that are standard for DNA analysis and data interpretation?

Objective Proof for the Finding:

The laboratory provided "Work Instructions" for the Biomek® 2000 robotic workstation using the Promega DNA IQ™ extraction chemistry (approved 7/28/2009) and the Quantifiler® Duo DNA Quantification Kit (approved 8/27/2009) but did not provide analytical procedures for these methods. Refer to finding under 9.1.1.

Standard 9.2 Does the laboratory use reagents that are suitable for the methods employed?

Note:

To successfully satisfy Standard 9.2, compliance must be demonstrated with all of the subcategories of Standard 9.2. There is no separate finding against this standard, only that the laboratory was not in compliance for Standard 9.2.2, and 9.2.2.b.

Standard 9.2.2 Are commercial reagents labeled with:

Standard 9.2.2.b The expiration date as provided by the manufacturer or as determined by the laboratory?

Objective Proof for the Finding:

There were several chemicals (DTT, Ethanol, Phenol, Sulfuric Acid) identified during the audit in both the nuclear and mtDNA laboratories that were not labeled with expiration dates provided by the manufacturer or that were determined by the laboratory.

Standard 9.3 Critical reagents shall include, but are not limited to, the reagents listed in Standards 9.3.1 and 9.3.2.

Standard 9.3.a Has the laboratory identified critical reagents?

Standard 9.3.b Has the laboratory evaluated critical reagents prior to use in casework?

Objective Proof for the Finding:

The list of Critical Reagents documented in the 2008 Standard Operating Procedures Section 8 Appendices (Revised 12-01-2008) Appendix V - 8.5.3 does not list the Quantifiler® Duo DNA Quantification Kit as a critical reagent. The laboratory procedures do not include written procedures detailing the acceptable range of results, procedures for addressing unacceptable data, and mechanisms used for documentation and subsequent approval/rejection of quality control data of the defined critical reagents. The Quantifiler® Duo kit was approved for casework on 8-27-09 but was not identified as a critical reagent until the approval of DNASOP-6, section 5.3 on 4-1-11. Through the evaluation of case files, the Quantifiler® Duo kit was in use on casework between the period of 8-27-09 and 4-1-11.

Appendix A: Findings and Responses

Findings:

Standard 9.3.1 Has the laboratory identified and evaluated the following:

Standard 9.3.1.a Test kits or systems for performing quantitative PCR?

Objective Proof for the Finding:

Although the laboratory has evaluated and implemented the use of the Quantifiler® Duo kit, it was not identified as a critical reagent in the 2008 Standard Operating Procedures Section 8 Appendices (Revised 12-01-2008) Appendix V - 8.5.3. which was in place from 8-27-09 through 4-1-11.

The laboratory routinely used the Quantifiler® Duo kit during the time frame noted above; however a procedure was not put into place until April 2011.

Standard 9.5 Does the laboratory monitor the analytical procedures using appropriate controls and standards?

Objective Proof for the Finding:

Differential extraction blanks are not processed concurrently and in parallel through the differential extraction process with the requisite evidence samples. There is one reagent blank processed per differential extraction set and the reagent blank is not split into two different fractions during processing and is therefore not being treated the same as the evidentiary items. The differential extraction blanks are therefore not adequately monitoring the analytical procedures used in the differential extraction process.

Standard 9.5.3 Are reagent blank controls associated with each extraction set being analyzed as follows:

Standard 9.5.3.1 Extracted concurrently?

Objective Proof for the Finding:

The discussion under this standard states, 'A laboratory shall associate a reagent blank control with each extraction set or batch of samples as defined by the laboratory. This control is treated the same as, and parallel to, the forensic and/or casework reference samples being analyzed.' Differential extraction blanks are not processed concurrently and in parallel through the differential extraction process with the requisite evidence samples. There is one reagent blank processed per differential extraction set and the reagent blank is not split into two different fractions during processing and is therefore not being treated the same as the evidentiary items.

Appendix A: Findings and Responses

Findings:

Standard 9.5.5 Does the laboratory check its DNA procedures either annually or whenever substantial changes are made to a procedure against an appropriate and available NIST standard reference material (SRM) or standard traceable to a NIST standard?

Objective Proof for the Finding:

The laboratory did not run a NIST traceable sample for autosomal and Y STR kits in 2009, 2010, or 2011. The laboratory used a sample, RKO (lot # 2-24-09) in 2009 and 2010 and RKO (lot # 3-28-11) in 2011 as the NIST traceable standards. These RKO samples were not directly traceable to the NIST SRM 2391b and/or 2395. The laboratory traced these RKO samples to previous RKO lots that were directly traceable to the NIST SRM 2391b and/or 2395; however in doing so, they are not direct NIST traceable samples.

Standard 9.6 Does the laboratory have and follow written guidelines for the interpretation of data?

Objective Proof for the Finding:

The laboratory's Section 7.0 PCR: Interpretation of Results states that sample mixture statistics will be done using the probability of inclusion (CPI) calculation using the following formula: $CPI = (PI)^1(PI)^2(PI)^3\dots(PI)^n$

The manual gives no guidance regarding the inclusion/exclusion of loci depending upon allelic activity under the reporting threshold. Numerous reports were reviewed where loci with alleles below the reporting threshold were used for statistical purposes to support an inclusion.

The laboratory's analytical threshold is 50 rfu and the reporting threshold is 75 rfu. Analysts mark alleles that fall between 50 and 75 rfu with an '*'. It is standard procedure to include loci that have alleles marked with an asterisk in the CPI calculation. For example, if a locus shows a 10 and 14 allele over 75 rfu and notes an 8 and 15 allele (each with an asterisk) that fall between 50 and 75 rfu, that locus will still be used in the CPI using only the 10 and 14 alleles. In these instances, the contributors having genotypes with an 8 and/or a 15 allele will not be encompassed by the interpreted alleles and are not being included as possible contributors which is a misrepresentation of the analytical data and makes the probability of inclusion less conservative. It should be noted that in the cases reviewed there were no major/minor contributors and many of the mixtures were indistinguishable. Additionally, it is unclear how the laboratory would handle an instance if the probative DNA profile alleles fell between 50 and 75 rfu, specifically would the locus be left out of the calculation or would an exclusion be declared.

Appendix A: Findings and Responses

Findings:

Standard 9.6.4 Does the laboratory have and follow documented procedures for mixture interpretation to include the following:

Standard 9.6.4.a Major and minor contributors?

Standard 9.6.4.b Inclusions and exclusions?

Standard 9.6.4.c Policies for reporting results and statistics?

Objective Proof for the Finding:

The laboratory has an Interpretation of Results SOP (Document ID: DNA SOP-5); however there are no documented procedures which specifically addresses how to discern major and minor contributors, inclusions and exclusions, and policies for reporting results and applicable statistics.

Standard 9.7 Does the laboratory have and follow a documented policy for detecting and controlling contamination?

Objective Proof for the Finding:

The laboratory does not have policy for detecting contamination. There is no documentation that describes whether peaks above and/or below the laboratory's analytical threshold are assessed as potential contamination. The audit team reviewed numerous case files where negative controls and/or reagent blanks showed allelic activity both above and/or below the laboratory's analytical threshold (50 and 75 rfu, respectively).

Additionally, as noted above, there is one reagent blank processed per differential extraction set and the reagent blank is not split into two different fractions during processing and is therefore not being treated the same as the evidentiary items and should a contamination event occur during this process it may go undetected without the appropriate use of this differential extraction reagent blank.

Standard 10.2 Does the laboratory have and follow a documented program for conducting performance checks and calibrating equipment and instruments?

Note:

To successfully satisfy Standard 10.2, compliance must be demonstrated with all of the subcategories of Standard 10.2. There is no separate finding against this standard, only that the laboratory was not in compliance for Standards 10.2.1, 10.2.1.3, 10.2.1.5, 10.2.1.6, and 10.2.1.7.

Appendix A: Findings and Responses

Findings:

Standard 10.2.1 At a minimum, are the following critical instruments or equipment performance-checked at least annually:

Standard 10.2.1.3 Thermal cycler temperature-verification system?

Objective Proof for the Finding:

The laboratory routinely does a calibration of the thermal cycler temperature-verification system; however the laboratory could not provide a written protocol defining when or how the calibration of the thermal cycler temperature-verification system is performed.

Standard 10.2.1 At a minimum, are the following critical instruments or equipment performance-checked at least annually:

Standard 10.2.1.5 Electrophoresis detection systems?

Objective Proof for the Finding:

The laboratory has two Agilent 2100 Bioanalyzers. There has been no annual performance checks conducted on either in 2008, 2009, and 2010. Staff interviewed stated that management said that there were insufficient funds available to carry this out.

Standard 10.2.1 At a minimum, are the following critical instruments or equipment performance-checked at least annually:

Standard 10.2.1.6 Robotic systems?

Objective Proof for the Finding:

The laboratory routinely performs a calibration of the Biomek® robotic workstation XY alignment after cleaning and preventive maintenance. The laboratory could not provide a written protocol defining the minimum requirements for the performance check or when the alignment of the XY axis of the Biomek® robotic workstation is performed.

Standard 10.2.1 At a minimum, are the following critical instruments or equipment performance-checked at least annually:

Standard 10.2.1.7 Genetic analyzers?

Objective Proof for the Finding:

Performance checks were conducted using the Identifiler® kit on the AB 3130xl Instrument #3 (Technical Leader approved 5/8/2008) and 3130xl #4 (Technical Leader approved 4/21/2010). No performance check was conducted using the PowerPlex® Y kit on either of these instruments. (refer to standard 8.6)

Appendix A: Findings and Responses

Findings:

Standard 10.3 Does the laboratory have a schedule and follow a documented program to ensure that instruments and equipment are maintained properly?

Objective Proof for the Finding:

The laboratory did not provide a written protocol defining when the maintenance of the Biomek® robotic workstation is performed.

Additionally, the two Agilent 2100 Bioanalyzers did not have any written performance check procedures.

Standard 10.4 Does the laboratory performance check new critical instruments and equipment, or critical instruments and equipment that have undergone repair, service or calibration, before their use in casework analysis?

Note:

To successfully satisfy Standard 10.4, compliance must be demonstrated with all of the subcategories of Standard 10.4. There is no separate finding against this standard, only that the laboratory was not in compliance for Standard 10.4.1.3.

Standard 10.4.1 At a minimum, are the following critical instruments or equipment performance-checked following repair, service, or calibration:

Standard 10.4.1.3 Genetic analyzers?

Objective Proof for the Finding:

The AB 3130xl instrument (serial # 1606-006) used for mtDNA analysis had performance maintenance conducted on 03/01/08, 03/31/09 and 03/29/10; however, no performance checks were conducted after any maintenance.

Standard 11.1 a. Does the laboratory have and follow written procedures for taking and maintaining case notes to support the conclusions drawn in laboratory reports?

Objective Proof for the Finding:

The laboratory does not always issue reports after conducting a DNA analysis on a case work that results in a 'hit' in the database.

Appendix A: Findings and Responses

Findings:

Continued from previous page:

Standard 11.1 a. Does the laboratory have and follow written procedures for taking and maintaining case notes to support the conclusions drawn in laboratory reports?

Objective Proof for the Finding - continued:

For example:

If the laboratory works a no suspect sexual a assault case and a male DNA profile is developed from one of the samples, this male profile will be uploaded into the database. If there is a 'hit' to this male DNA profile, the laboratory will issue a hit letter stating that a report will not be issued until a comparison standard is submitted. If no standard is submitted, no report regarding the DNA analysis on this sample, or an other samples processed in this case, will ever be issued by the laboratory. Alternatively, if the laboratory works a no suspect sexual a assault case and a male DNA profile is developed from one of the samples, this male profile will be uploaded into the database. If there is NO 'hit' to this male DNA profile, the laboratory will issue a report regarding the DNA analysis of all samples processed in the case.

There is no written procedures outlining this practice. If comparison samples are not issued, then there is no documentation that these cases were ever analyzed by the laboratory which could adversely impact investigations. (see standard 12.2)

Standard 11.2 Do the laboratory reports include the following elements:

Standard 11.2.9 Signature and title, or equivalent identification, of the person accepting responsibility for the content of the report?

Objective Proof for the Finding:

The discussion under this standard states, 'One person shall accept responsibility for the content of the report.' Each laboratory report reviewed by the audit team contained the names, signatures and titles of two forensic science examiners. The laboratory does not distinguish between the two examiner's roles in the signature sections of the report. Therefore the analyst accepting responsibility for the content of the laboratory report is not identified. Both examiners are listed as "Forensic Science Examiner" in the reports.

The Technical Leader, in interviews, stated that the person signing on the left side of the report was the individual accepting responsibility for the report. The Quality Manual, Section 3, revision 4 (4/2007) - 3.1 General Considerations, section B states, 'Each original Examination Report must be signed by the primary case examiner and a reviewing examiner as "co-signer.'" The quality system is silent on this issue as to which examiner signs on the left or right side of the report is taking responsibility.

Appendix A: Findings and Responses

Findings:

Continued from previous page:

Standard 11.2 Do the laboratory reports include the following elements:

Standard 11.2.9 Signature and title, or equivalent identification, of the person accepting responsibility for the content of the report?

Objective Proof for the Finding - continued:

One analyst had performed DNA analysis on a criminal paternity case and when interviewed regarding that statistical application of paternity statistics he said he was not comfortable with conducting the statistical analysis or with testifying to them and did not recall being trained in paternity statistics. When this was brought to the Technical Leader's attention by the audit team, he stated that the analyst would not be allowed to testify to the paternity statistics and that they would have the other analyst that signed the report (on the right side) testify instead (see findings under standard 5.1.2.2). The Technical Leader stated that both analysts were assuming responsibility for the content of the report; however later discussions indicated that this analyst was the technical reviewer. Therefore it appears that both analysts signing the report are assuming technical responsibility for the report.

Standard 12.1 Does the laboratory conduct and document administrative and technical reviews of all case files and reports to ensure that conclusions and supporting data are reasonable and within the constraints of scientific knowledge?

Objective Proof for the Finding:

The Quality Manual, Section 4, revision 4 (4/2007) - 4.2 Case Review Procedures 4.2.1 Administrative & Technical Reviews, section 1, states, 'All cases shall be reviewed by a second examiner and/or immediate supervisor. The second examiner or supervisor will act as co-signer of the examination report.' The discussion under standard 12.1 states that the technical reviewer is an employee who is a current or previously qualified analyst in the methodology being reviewed that performs a technical review of, and is not an author of, the applicable report or its contents.' It is unclear to the audit team the specific role of each analyst signing the reports and appears that there may be dual roles as far as responsibility of the contents of the report which means that the second signer should not be technically reviewing the case file.

Appendix A: Findings and Responses

Findings:

Standard 12.2 Does the laboratory document the completion of the technical review of forensic casework, and does it include the following elements:

Standard 12.2.1 A review of all case notes, worksheets, and electronic data (or printed electropherograms/images) that support the conclusions?

Standard 12.2.2 A review of all DNA types to verify that they are supported by the raw or analyzed data (electropherograms or images)?

Standard 12.2.3 A review of all profiles to verify correct inclusions and exclusions (if applicable) as well as a review of any inconclusive result for compliance with laboratory guidelines?

Standard 12.2.4 A review of all controls, internal lane standards, and allelic ladders to verify that the expected results

Standard 12.2.5 A review of statistical analysis, if applicable?

Objective Proof for the Finding:

The Laboratory's DNA QA/QC Case Work Checklist Review sheet (QR-4) does not include the following elements:

1. A review of all case notes, worksheets, and electronic data (or printed electropherograms/images).
2. A review of all DNA types.
3. A review of all profiles to verify correct inclusions and exclusions (if applicable) as well as a review of any inconclusive result for compliance with laboratory guidelines.
4. A review of all controls, internal lane standards, and allelic ladders.
5. A review of the statistical analysis.

Note: the mtDNA technical review sheet does cover standards 12.2.4.2 and 12.2.4.4; but not 12.2.4.2, 12.2.4.3, or 12.2.4.5.

Standard 12.2.6 A review of the final report to verify that the results/conclusions are supported by the data?

Standard 12.2.6.a Does the report address each tested item or its probative fraction?

Objective Proof for the Finding:

There were several cases reviewed by the audit team in which laboratory reports on forensic casework were issued which did not address each item tested or its probative fraction.

Appendix A: Findings and Responses

Findings:

Continued from previous page:

Standard 12.2.6 A review of the final report to verify that the results/conclusions are supported by the data?

Standard 12.2.6.a Does the report address each tested item or its probative fraction?

Objective Proof for the Finding - continued:

The discussion under this standard states, 'Final reports of forensic casework shall address each tested item or its probative fraction. Any stain, sample, or item on which an attempt is made to isolate DNA, regardless of the outcome or result, must be addressed in the final report.' This implies that all forensic casework conducted by a laboratory should result in a report addressing each tested item. The laboratory does not always issue reports after conducting a DNA analysis on a case work that results in a 'hit' in the database. Please refer to the finding under standard 11.1.a for specific details.

Standard 12.2.7.2 Prior to entry of a DNA profile into a searchable category of SDIS, were the following criteria verified by two concordant assessments by a qualified analyst or technical reviewer?

Standard 12.2.7.2.a Eligibility for CODIS?

Standard 12.2.7.2.b Correct DNA types?

Standard 12.2.7.2.c Appropriate specimen category?

Objective Proof for the Finding:

All case files reviewed by the audit team which contained CODIS entries revealed the 'marked by date' (i.e. date uploaded into SDIS) was performed prior to any concordant assessments by a qualified analyst or technical reviewer.

The mtDNA case files/profiles were technically reviewed prior to upload; however, eligibility and category were NOT checked prior to entry.

Section 1.0 of the General Procedures for the Analysis of mtDNA Evidence - subsection 1.12 Case review, states, "All mtDNA cases/reports are technically reviewed by at least one mtDNA analyst in addition to the primary examiner. All mtDNA cases/reports are administratively reviewed by the mtDNA Technical Leader. mtDNA sequences must be second analyzed by at least one qualified mtDNA examiner. In the event that there is only one qualified mtDNA analyst available for an extended period of time, the documents, worksheets and data for each case completed are sent to one of the other three mtDNA Regional Laboratories for second analysis and technical review." This laboratory went off line as an FBI regional laboratory in 2009 and this subsection was not updated until the April 2011 revision.

Appendix A: Findings and Responses

Findings:

Standard 12.4 Does the laboratory document the elements of a technical and administrative review?

Standard 12.4.a Are case files reviewed and documented according to the laboratory's procedures?

Objective Proof for the Finding:

The laboratory manuals/procedures do not document what constitutes technical and administrative reviews as specified in these standards.

Standard 13.1.7 Does the laboratory include, at a minimum, the following criteria for evaluating proficiency test results:

Standard 13.1.7.1 Evaluation:

Standard 13.1.7.1.a Are all reported inclusions correct?

Standard 13.1.7.1.b Are all reported exclusions correct?

Standard 13.1.7.1.c Are all reported genotypes and/or phenotypes correct or incorrect according to consensus results or within the laboratory's interpretation guidelines?

Objective Proof for the Finding:

The Quality Manual, Section 9.0 Proficiency Testing Procedures (Version 4, Revised 04/2007) does not state that the review of the DNA proficiency tests will be evaluated to confirm that all reported inclusions (if applicable) are correct, all reported exclusions (if applicable) are correct, or that all reported genotypes and/or phenotypes are correct or incorrect according to consensus results or within the laboratory's interpretation guidelines.

Appendix A: Findings and Responses

Findings:

Standard 14.1 For a corrective action plan:

Standard 14.1.a Has the laboratory established and followed a corrective action plan that addresses discrepancies detected in proficiency tests and casework analysis?

Standard 14.1.b Does the corrective action plan, at a minimum, address the following:

Standard 14.1.b.1 Define what level/type of discrepancies are applicable to this practice?

Standard 14.1.b.2 Identify (when possible) the cause of the discrepancy?

Standard 14.1.b.3 Effect of the discrepancy?

Standard 14.1.b.4 Corrective actions taken?

Standard 14.1.b.5 Preventative measures taken (where applicable) to minimize its reoccurrence?

Standard 14.1.b.6 Is documentation of all corrective actions maintained in accordance with Standard 3.2?

Objective Proof for the Finding:

The laboratory has a corrective action plan (QA Program, Section 8.1-Rev. 4); however the plan does not address discrepancies detected in casework analysis. In addition, the plan does not define level/type of discrepancies, identify causes and effects of discrepancies, corrective actions, preventative measures taken to minimize reoccurrences or if the documentation of all corrective actions maintained in accordance with Standard 3.2. There is also no documentation to show that they are doing this in practice.

The laboratory does not define when an incident will be handled using the corrective action plan outlined in the QA Program.

For example, as of December 2010, the laboratory was out of compliance with the QAS because they did not have an external audit. This was noted in the internal audit they conducted in 2010; however they did not start a corrective action report until June 2011. The laboratory did not have any other documented corrective actions since the December 2008 external audit.

Appendix A: Findings and Responses

Findings:

Standard 15.1 Has the laboratory been audited annually in accordance with the FBI DNA Quality Assurance Standards? (first section)

Standard 15.2 Has an external audit been conducted at least once every two years?

Standard 15.2.a By a qualified auditor?

Standard 15.2.b By a current or previously qualified analyst in the laboratory's current DNA technologies and platform?

Objective Proof for the Finding:

The laboratory's Quality Manual, Section 8.2.1 (d) Rev. 4 states "The DNA Unit procedures shall also be reviewed by an external auditor approximately once every two years (minimum), in accordance with FBI NDIS/CODIS requirements". The laboratory had their last external audit in December 2008. There was no documentation to demonstrate the laboratory had an external audit in 2010. Therefore standards 15.2.a and 15.2.b are also marked NO.

Standard 15.3 For internal audits, has the laboratory maintained documentation that the auditor(s) for this inspection include:

Standard 15.3.b A current or previously qualified analyst in the laboratory's current DNA technologies and platform?

Objective Proof for the Finding:

The laboratory did not complete Appendix C of the 2009 FBI Quality Assurance Standards Audit Document related to the internal audits conducted in 2009 or 2010. The audit team was unable to determine if a current or previously qualified analyst in the laboratory's current DNA technologies and platforms were included on the internal audit team. The signature page of internal audit reports, reviewed by the audit team, listed audit team members. No member of the internal audit teams were or had been a previously qualified mtDNA analyst.


Standard 16.1 Does the laboratory have and follow a documented environmental health and safety program that includes, at a minimum, the following:

Objective Proof for the Finding:

The laboratory's Safety Manual, Section 5.8A (Rev. 3) states "The Laboratory Safety Officer will meet with the safety committee on a quarterly basis to discuss any non-emergency issues/concerns that have risen since the previous meeting". According to the Quality Manager/Safety Officer, there have never been any safety committee meetings.

Appendix A: Findings and Responses

Responses:

A large, empty rectangular box with a thin black border, occupying most of the page below the 'Responses:' header. It is intended for the user to provide their responses to the findings listed in the adjacent column.

Appendix C – Auditor Self-Certification for QAS Audits

Section 1 – to be completed by the laboratory being audited:

Laboratory being audited: CT State Div. of Scientific Services Laboratories-CW As of [date] June 28, 2011

Technologies currently in use: Identifiler, PowerPlex Y, mtDNA

Platforms currently in use: ABI 3130xl

Validations needing to be memorialized: Quantifiler Human + Y, Duo; DNA, IQ on Biomek 2000, BSD Duet

Outsourcing agreements in place or in process: NA

The laboratory being audited may request documentation for the information reported in Section 2 below.

Section 2 – to be completed by the auditor who will sign the attestation statement below the questions and (a) for external audits, return to the laboratory prior to the scheduled audit date; or (b) for internal audits, maintain in the laboratory's files.

Auditor Qualifications:

Name of Auditor: Sylvia Thurmond

Auditor's Employer: Denver Police Dept Crime Lab

Auditor's Title or Position: Forensic Scientist II - DNA Analyst

Qualified Auditor¹: Yes No (Check One)

Year Completed FBI DNA Auditor Class: 2007

Current or Previously Qualified DNA Analyst: Yes No (Check One)

Current or Previously Qualified in Casework, Database Analysis, or Both²:

Casework Database Both (Check One) - Current: Casework Previous: Database

Technologies Currently or Previously Qualified In (e.g., STR, mtDNA) (Please List):

Current: STR / Y-STR Previous: mtDNA

Platforms Currently or Previously Qualified In (e.g., Gel based/CE) (Please List):

Current: CE, Robotics (EZ1, Quasymphony, Qiagility) Previous: Gel

I verify that:

I understand the requirements of Standard 15.2³; and

I have no conflicts of interest with the laboratory being audited; and

The information contained in Section 2 above is correct.

¹ A Qualified Auditor is a current or previously qualified DNA analyst who has successfully completed the FBI DNA Auditor training course.

² If the laboratory being audited performs both casework and database analyses, then the audit team or auditor must be qualified in both casework and database analyses.

³ Standard 15.2 requires that "at least once every two years, an external audit shall be conducted by an audit team comprised of qualified auditors from a second agency(ies) and having at least one team member who is or has been previously qualified in the laboratory's current DNA technologies and platform."

Signed By  Date 6/30/11

Appendix C – Auditor Self-Certification for QAS Audits

Section 1 – to be completed by the laboratory being audited:

Laboratory being audited: CT State Div. of Scientific Services Laboratories-CW As of [date] June 28, 2011

Technologies currently in use: Identifiler, PowerPlex Y, mtDNA

Platforms currently in use: ABI 3130xl

Validations needing to be memorialized: Quantifiler Human + Y, Duo; DNA, IQ on Biomek 2000, BSD Duet

Outsourcing agreements in place or in process: NA

The laboratory being audited may request documentation for the information reported in Section 2 below.

Section 2 – to be completed by the auditor who will sign the attestation statement below the questions and (a) for external audits, return to the laboratory prior to the scheduled audit date; or (b) for internal audits, maintain in the laboratory's files.

Auditor Qualifications:

Name of Auditor: Christie Smith

Auditor's Employer: AZ DPS Crime Lab

Auditor's Title or Position: Criminalist II

Qualified Auditor¹: Yes No (Check One)

Year Completed FBI DNA Auditor Class: 2005

Current or Previously Qualified DNA Analyst: Yes No (Check One)

Current or Previously Qualified in Casework, Database Analysis, or Both²:

Casework Database Both (Check One)

Technologies Currently or Previously Qualified In (e.g., STR, mtDNA) (Please List):

mtDNA

Platforms Currently or Previously Qualified In (e.g., Gel based/CE) (Please List):

CE

I verify that:

I understand the requirements of Standard 15.2³; and

I have no conflicts of interest with the laboratory being audited; and

The information contained in Section 2 above is correct.

¹ A Qualified Auditor is a current or previously qualified DNA analyst who has successfully completed the FBI DNA Auditor training course.

² If the laboratory being audited performs both casework and database analyses, then the audit team or auditor must be qualified in both casework and database analyses.

³ Standard 15.2 requires that "at least once every two years, an external audit shall be conducted by an audit team comprised of qualified auditors from a second agency(ies) and having at least one team member who is or has been previously qualified in the laboratory's current DNA technologies and platform."

Signed By Christie Smith Date 7/1/11

Appendix C – Auditor Self-Certification for QAS Audits

Section 1 – to be completed by the laboratory being audited:

Laboratory being audited: CT State Div. of Scientific Services Laboratories-CW As of [date] June 28, 2011

Technologies currently in use: Identifiler, PowerPlex Y, mtDNA

Platforms currently in use: ABI 3130xl

Validations needing to be memorialized: Quantifiler Human + Y, Duo; DNA, IQ on Biomek 2000, BSD Duet

Outsourcing agreements in place or in process: NA

The laboratory being audited may request documentation for the information reported in Section 2 below.

Section 2 – to be completed by the auditor who will sign the attestation statement below the questions and (a) for external audits, return to the laboratory prior to the scheduled audit date; or (b) for internal audits, maintain in the laboratory's files.

Auditor Qualifications:

Name of Auditor: Jason C. Hodges

Auditor's Employer: WV State Police Forensic Lab

Auditor's Title or Position: Forensic Analyst/DNA QA Officer

Qualified Auditor¹: Yes No (Check One)

Year Completed FBI DNA Auditor Class: 2007

Current or Previously Qualified DNA Analyst: Yes No (Check One)

Current or Previously Qualified in Casework, Database Analysis, or Both²:

Casework Database Both (Check One)

Technologies Currently or Previously Qualified In (e.g., STR, mtDNA) (Please List):

STR's: PP16 and PPY

Platforms Currently or Previously Qualified In (e.g., Gel based/CE) (Please List):

CE: 3130/3130xl

I verify that:

I understand the requirements of Standard 15.2³ ; and

I have no conflicts of interest with the laboratory being audited; and


The information contained in Section 2 above is correct.

¹ A Qualified Auditor is a current or previously qualified DNA analyst who has successfully completed the FBI DNA Auditor training course.

² If the laboratory being audited performs both casework and database analyses, then the audit team or auditor must be qualified in both casework and database analyses.

³ Standard 15.2 requires that "at least once every two years, an external audit shall be conducted by an audit team comprised of qualified auditors from a second agency(ies) and having at least one team member who is or has been previously qualified in the laboratory's current DNA technologies and platform."

Jason C. Hodges 6/29/11

Signed By  Date 6/29/11

Appendix C – Auditor Self-Certification for QAS Audits

Section 1 – to be completed by the laboratory being audited:

Laboratory being audited: CT State Div. of Scientific Services
Laboratories-CW As of [date] June 28, 2011

Technologies currently in use: Identifiler, PowerPlex Y, mtDNA

Platforms currently in use: ABI 3130xl

Validations needing to be memorialized: Quantifiler Human + Y, Duo; DNA, IQ on Biomek 2000, BSD Duet

Outsourcing agreements in place or in process: NA

The laboratory being audited may request documentation for the information reported in Section 2 below.

Section 2 – to be completed by the auditor who will sign the attestation statement below the questions and (a) for external audits, return to the laboratory prior to the scheduled audit date; or (b) for internal audits, maintain in the laboratory's files.

Auditor Qualifications:

Name of Auditor: Lonnie Ginsberg

Auditor's Employer: AL Dept. of Forensic Sciences

Auditor's Title or Position: Lab Director

Qualified Auditor¹: Yes No (Check One)

Year Completed FBI DNA Auditor Class: 2001, 2003 (refresher) 2004, 2009

Current or Previously Qualified DNA Analyst: Yes No (Check One)

Current or Previously Qualified in Casework, Database Analysis, or Both²:

Casework Database Both (Check One)

Technologies Currently or Previously Qualified In (e.g., STR, mtDNA) (Please List):

STR, Y-STR

Platforms Currently or Previously Qualified In (e.g., Gel based/CE) (Please List):

CE

I verify that:

I understand the requirements of Standard 15.2³ ; and

I have no conflicts of interest with the laboratory being audited; and

The information contained in Section 2 above is correct.

¹ A Qualified Auditor is a current or previously qualified DNA analyst who has successfully completed the FBI DNA Auditor training course.

² If the laboratory being audited performs both casework and database analyses, then the audit team or auditor must be qualified in both casework and database analyses.

³ Standard 15.2 requires that "at least once every two years, an external audit shall be conducted by an audit team comprised of qualified auditors from a second agency(ies) and having at least one team member who is or has been previously qualified in the laboratory's current DNA technologies and platform."

Signed By Liamie Gensby Date 6/29/11

Appendix C – Auditor Self-Certification for QAS Audits

Section 1 – to be completed by the laboratory being audited:

Laboratory being audited: CT State Div. of Scientific Services Laboratories-CW As of [date] June 28, 2011

Technologies currently in use: Identifiler, PowerPlex Y, mtDNA

Platforms currently in use: ABI 3130xl

Validations needing to be memorialized: Quantifiler Human + Y, Duo; DNA, IQ on Biomek 2000, BSD Duet

Outsourcing agreements in place or in process: NA

The laboratory being audited may request documentation for the information reported in Section 2 below.

Section 2 – to be completed by the auditor who will sign the attestation statement below the questions and (a) for external audits, return to the laboratory prior to the scheduled audit date; or (b) for internal audits, maintain in the laboratory's files.

Auditor Qualifications:

Name of Auditor: LISA GEFRIDES

Auditor's Employer: SELF

Auditor's Title or Position: CONSULTANT

Qualified Auditor¹: Yes No (Check One)

Year Completed FBI DNA Auditor Class: Refresher in 2009

Current or Previously Qualified DNA Analyst: Yes No (Check One)

Current or Previously Qualified in Casework, Database Analysis, or Both²:

Casework Database Both (Check One)

Technologies Currently or Previously Qualified In (e.g., STR, mtDNA) (Please List):

STR, Y-STR, mtDNA

Platforms Currently or Previously Qualified In (e.g., Gel based/CE) (Please List):

371 (mtDNA) 310, 3100 (STR) 3130XL (STR, Y-STR)

I verify that:

I understand the requirements of Standard 15.2³; and

I have no conflicts of interest with the laboratory being audited; and

The information contained in Section 2 above is correct.

¹ A Qualified Auditor is a current or previously qualified DNA analyst who has successfully completed the FBI DNA Auditor training course.

² If the laboratory being audited performs both casework and database analyses, then the audit team or auditor must be qualified in both casework and database analyses.

³ Standard 15.2 requires that "at least once every two years, an external audit shall be conducted by an audit team comprised of qualified auditors from a second agency(ies) and having at least one team member who is or has been previously qualified in the laboratory's current DNA technologies and platform."

Lisa Gefrides 063011

Appendix D – Personnel Meeting Minimum Education, Experience, and Training Qualifications As Assessed By External Audit

To be completed by the audit team.

In accordance with Standards 15.1 and 15.2.1, this form shall be used to document the evaluation and approval of analysts, CODIS administrators and technical leaders during an external audit. Section 1 is for documenting personnel who have received two successive separate external audit approvals of their education, experience, and training qualifications. Section 1 should be used to document all individuals who have received two successive separate audit approvals of their education, experience, and training qualifications, regardless of whether the individual is still employed by the laboratory. The date of the prior audit approvals should be noted in this Section, when known.

Section 2 is for documenting personnel who are receiving the **first** external audit approval of their education, experience, and training qualifications.

Section 1 documents those personnel who have received two successive external audit approvals of their education, experience, and training qualifications.

**Section 1. (a) – Approvals Between July 1, 2004 and June 30, 2009
Laboratory personnel who have been evaluated after July 1, 2004, and approved under two successive, separate external audits as meeting the education, experience, and training qualifications required under Standard 5.1 of the 1998 Quality Assurance Standards for Forensic DNA Testing Laboratories:**

Analyst(s):

Michael Bourke (3/2005 & 6/2006) –technical review/report writing only (autosomal STR and YSTR casework)
Eric J. Carita (6/2006 & 12/2008) – casework (autosomal STR and YSTR)
Patricia Johannes (3/2005 & 6/2006) – casework (autosomal STR and YSTR)
Kristen Madel (6/2006 & 12/2008) – mtDNA
Christine Roy (3/2005 & 6/2006) – casework (autosomal STR and YSTR)
Kristin Sasinouski (6/2006 & 12/2008) - mtDNA
John Schienman (6/2006 & 12/2008) – casework (autosomal STR and YSTR)
Nicholas Yang (3/2005 & 6/2006) – casework (autosomal STR and YSTR)
Carl Ladd (3/2005 & 6/2006) - technical review/report writing only

Approvals after July 1, 2004 and this audit makes two successive, separate external audits:

Steven Bryant (12/2008 & 7/2011) – casework (autosomal STR and YSTR)
Heather Degnan (12/2008 & 7/2011) – casework (autosomal STR and YSTR)
Angela Przech (12/2008 & 7/2011) – casework (autosomal STR and YSTR)
Melanie Russell (formally Ktorides) - (12/2008 & 7/2011) - casework (autosomal STR and YSTR)

Technical Leader(s):

Carl Ladd (6/2006 & 12/2008) - nuclear DNA Technical Leader

Section 1. (b) – Approvals After July 1, 2009 Laboratory personnel who have been evaluated after July 1, 2009, and approved under two successive, separate external audits as meeting the education, experience, and training qualifications required under Standard 5.1 of the 2009 Quality Assurance Standards for DNA Databasing Laboratories:

Analyst(s):

Casework CODIS Administrator(s):

Technical Leader(s):

Section 2. (a) – For Personnel Appointed or Hired Prior to July 1, 2009

Laboratory personnel who were appointed or hired prior to July 1, 2009, and approved for the first time as meeting the education, experience, and training qualifications required under Standard 5.1 of the 1998 Quality Assurance Standards for Forensic DNA Testing Laboratories:

Analyst(s):

Jillian Echard (formerly Liotti) - was listed as an analyst in the 2006 ASCLD/LAB audit. She was working as a technician during the time span and should not have been listed as an analyst. She is now working in an analyst capacity – casework (autosomal STR and YSTR)

Adrienne Schoefer (formerly Richards) - was listed as an analyst in the 2006 ASCLD/LAB audit. She was working as a technician during the time span and should not have been listed as an analyst. She is now working in an analyst capacity– mtDNA

Technical Leader(s):

Empty box for Technical Leader(s) information.

Section 2. (b) – For Personnel Appointed or Hired On or After July 1, 2009 Laboratory personnel who have been evaluated after July 1, 2009, and approved for the first time as meeting the education, experience, and training qualifications required under Standard 5.1 of the 2009 Quality Assurance Standards for DNA Databasing Laboratories:

Analyst(s):

Cheryl Civitello (7/2011) – casework (autosomal STR and YSTR)

Megan Devlin (7/2011) – casework (autosomal STR and YSTR)

Dahong Sun (7/2011) – casework (autosomal STR and YSTR)

Casework CODIS administrator(s):

Michael Bourke

Technical Leader(s):

Carl Ladd - mtDNA Technical Leader

Appendix E – Approved Validations

This form may be used to document the evaluation and approval of validations by the external audit team according to Standard 8; this documentation to be maintained by the audited laboratory to comply with Standard 15.2.2.

To be completed by the audit team:

List of validations, if any, evaluated and approved during this audit:

Qiagen DyeEx 2.0 Spin Column filtration system for mtDNA. Dye removal evaluation. This validation was performed in 2006; however was never captured in an external audit document. Reproducibility and precision were assessed.